

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: P3845 SAS No.: P3845 SDG NO.: P3845

EPA Sample No.: SSTDICCC0.4 Date Analyzed: 09/18/2024

Lab File ID: BN034035.D Time Analyzed: 13:05

Instrument ID: BNA_N GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	6856	7.437	19411	10.20	11055	14.08
UPPER LIMIT	13712	7.937	38822	10.697	22110	14.582
LOWER LIMIT	3428	6.937	9705.5	9.697	5527.5	13.582
EPA SAMPLE NO.						
01 PT-ACIDS-WP	7056	7.44	21248	10.19	11207	14.08
02 PT-ACIDS-WPDL	7922	7.44	22980	10.19	11429	14.08

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.