

6C

SEMIVOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: CHEM02
Lab Code: CHEM Case No.: P3845 SAS No.: P3845 SDG No.: P3845
Instrument ID: BNA_M Calibration Date(s): 09/20/2024 09/20/2024
Calibration Time(s): 09:40 14:17

LAB FILE ID:		RRF2.5 = BM047581.D		RRF005 = BM047582.D		RRF010 = BM047583.D		
		RRF020 = BM047584.D		RRF040 = BM047585.D		RRF050 = BM047586.D		
COMPOUND	RRF2.5	RRF005	RRF010	RRF020	RRF040	RRF050	RRF	% RSD
2-Fluorophenol		1.302	1.285	1.295	1.279	1.304	1.305	1.7
Phenol-d6		1.663	1.654	1.703	1.699	1.747	1.720	3.2
Nitrobenzene-d5		0.379	0.382	0.399	0.404	0.409	0.402	4.1
2-Fluorobiphenyl		1.390	1.325	1.382	1.363	1.419	1.413	5.0
2,4,6-Tribromophenol		0.158	0.162	0.176	0.187	0.194	0.185	11.3
Atrazine		0.157	0.155	0.156	0.162	0.130	0.144	12.5
Terphenyl-d14		1.016	0.961	1.070	1.128	1.211	1.115	9.3