

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: P3845 SAS No.: P3845 SDG NO.: P3845

EPA Sample No.: SSTDCCC040 Date Analyzed: 09/23/2024

Lab File ID: BM047598.D Time Analyzed: 09:42

Instrument ID: BNA_M GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	308343	7.845	1217070	10.64	646829	14.47
UPPER LIMIT	616686	8.345	2434140	11.139	1293660	14.974
LOWER LIMIT	154172	7.345	608535	10.139	323415	13.974
EPA SAMPLE NO.						
01 RR-TRIAZINE-WP	260539	7.84	961904	10.63	453032	14.47

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.