

Surrogate Summary

SDG No.: P3845

Client: Chemtech Consulting Group

Analytical Method: SW624.1

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
P3845-02	PT-VOA-WP	1,2-Dichloroethane-d4	30	31.2	104	91	110
		Toluene-d8	30	30.1	100	91	112
		4-Bromofluorobenzene	30	29.9	100	63	112
VN0909WBL02	VN0909WBL02	1,2-Dichloroethane-d4	30	29.8	99	91	110
		Toluene-d8	30	30.1	100	91	112
		4-Bromofluorobenzene	30	27.3	91	63	112
VN0909WBS01	VN0909WBS01	1,2-Dichloroethane-d4	30	30.0	100	91	110
		Toluene-d8	30	29.8	99	91	112
		4-Bromofluorobenzene	30	28.9	96	63	112
VN0909WBSD0	VN0909WBSD01	1,2-Dichloroethane-d4	30	29.6	99	91	110
		Toluene-d8	30	30.1	100	91	112
		4-Bromofluorobenzene	30	29.8	99	63	112