

Instrument ID: MSVOA_N

Daily Analysis Runlog For Sequence/QC Batch ID # VN091024

Review By	John Carlone	Review On	9/11/2024 4:18:14 PM
Supervise By	Semsettin Yesilyurt	Supervise On	9/11/2024 7:08:33 PM
SubDirectory	VN091024	HP Acquire Method	HP Processing Method 82N091024W.M

STD. NAME	STD REF.#
Tune/Reschk	VP130211,VP130225
Initial Calibration Stds	VP130230,VP130231,VP130232,VP130233,VP130234,VP130235
CCC	VP130212,VP130213,VP130226
Internal Standard/PEM	
ICV/I.BLK	VP130236
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VN083741.D	10 Sep 2024 08:02		JC\MD	Ok
2	VSTDICC100	VSTDICC100	VN083742.D	10 Sep 2024 09:08	Good for DOD & Non DOD	JC\MD	Ok,M
3	VSTDICCC050	VSTDICCC050	VN083743.D	10 Sep 2024 09:32	LR- 3, 18	JC\MD	Ok,M
4	VSTDICC020	VSTDICC020	VN083744.D	10 Sep 2024 09:56	Not used	JC\MD	Not Ok
5	VSTDICC010	VSTDICC010	VN083745.D	10 Sep 2024 10:20		JC\MD	Ok,M
6	VSTDICC005	VSTDICC005	VN083746.D	10 Sep 2024 10:44		JC\MD	Ok,M
7	VSTDICC001	VSTDICC001	VN083747.D	10 Sep 2024 11:32		JC\MD	Ok,M
8	IBLK	IBLK	VN083748.D	10 Sep 2024 11:56		JC\MD	Ok
9	VSTDICC020	VSTDICC020	VN083749.D	10 Sep 2024 12:43		JC\MD	Ok,M
10	VSTDICV050	ICVVN091024	VN083750.D	10 Sep 2024 13:36		JC\MD	Ok,M
11	VN0910MBL01	VN0910MBL01	VN083751.D	10 Sep 2024 14:10		JC\MD	Ok
12	VN0910WBL01	VN0910WBL01	VN083752.D	10 Sep 2024 14:34		JC\MD	Ok
13	VN0910WBS01	VN0910WBS01	VN083753.D	10 Sep 2024 15:34	BS low	JC\MD	Ok,M
14	VN0910WBSD01	VN0910WBSD01	VN083754.D	10 Sep 2024 16:07		JC\MD	Ok,M
15	P3845-01	PT-VOA-WP	VN083755.D	10 Sep 2024 16:31	BS low, Need 4X	JC\MD	Dilution
16	IBLK	IBLK	VN083756.D	10 Sep 2024 16:55		JC\MD	Ok
17	P3841-02DL	OW-07B-65.5-090424D	VN083757.D	10 Sep 2024 17:20	vial B pH<2	JC\MD	Ok
18	P3841-03DL	MW-11B-37.5-090424D	VN083758.D	10 Sep 2024 17:44	vial B pH<2	JC\MD	Ok

Instrument ID: MSVOA_N

Daily Analysis Runlog For Sequence/QC Batch ID # VN091024

Review By	John Carlone	Review On	9/11/2024 4:18:14 PM
Supervise By	Semsettin Yesilyurt	Supervise On	9/11/2024 7:08:33 PM
SubDirectory	VN091024	HP Acquire Method	HP Processing Method 82N091024W.M
STD. NAME	STD REF.#		
Tune/Reschk	VP130211,VP130225		
Initial Calibration Stds	VP130230,VP130231,VP130232,VP130233,VP130234,VP130235		
CCC	VP130212,VP130213,VP130226		
Internal Standard/PEM			
ICV/I.BLK	VP130236		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	P3868-01	OWBR-06-219-090524	VN083759.D	10 Sep 2024 18:08	vial B pH<2	JC\MD	Ok
20	P3868-02	OWBR-01-170-090524	VN083760.D	10 Sep 2024 18:32	vial B pH<2	JC\MD	Ok
21	P3907-02	PL-701-COMP-54	VN083761.D	10 Sep 2024 18:56	vial A pH#5.0	JC\MD	Ok,M
22	P3907-08	PL-701-COMP-55	VN083762.D	10 Sep 2024 19:20	vial A pH#5.0	JC\MD	Ok,M
23	VSTDCCC050	VSTDCCC050EC	VN083763.D	10 Sep 2024 19:44		JC\MD	Ok,M
24	BFB	BFB	VN083764.D	11 Sep 2024 02:03		JC\MD	Ok
25	VSTDCCC050	VSTDCCC050	VN083765.D	11 Sep 2024 02:27		JC\MD	Ok,M
26	VN0910MBL02	VN0910WBL03	VN083766.D	11 Sep 2024 02:51		JC\MD	Ok
27	VN0910WBL02	VN0910WBL04	VN083767.D	11 Sep 2024 03:15		JC\MD	Ok
28	VN0910WBS02	VN0910WBS02	VN083768.D	11 Sep 2024 03:39		JC\MD	Ok,M
29	VN0910WBSD02	VN0910WBSD02	VN083769.D	11 Sep 2024 04:03		JC\MD	Ok,M
30	PB163204TB	PB163204TB	VN083770.D	11 Sep 2024 04:27		JC\MD	Ok,M
31	PB163204ZHE#01	PB163204ZHE#01	VN083771.D	11 Sep 2024 04:50		JC\MD	Ok,M
32	PB163204ZHE#02	PB163204ZHE#02	VN083772.D	11 Sep 2024 05:14		JC\MD	Ok,M
33	PB163204ZHE#03	PB163204ZHE#03	VN083773.D	11 Sep 2024 05:38		JC\MD	Ok,M
34	PB163204ZHE#04	PB163204ZHE#04	VN083774.D	11 Sep 2024 06:02		JC\MD	Ok,M
35	PB163204ZHE#05	PB163204ZHE#05	VN083775.D	11 Sep 2024 06:26		JC\MD	Ok,M
36	PB163204ZHE#06	PB163204ZHE#06	VN083776.D	11 Sep 2024 06:50		JC\MD	Ok
37	PB163204ZHE#07	PB163204ZHE#07	VN083777.D	11 Sep 2024 07:14		JC\MD	Ok,M

Instrument ID: MSVOA_N

Daily Analysis Runlog For Sequence/QC Batch ID # VN091024

Review By	John Carlone	Review On	9/11/2024 4:18:14 PM
Supervise By	Semsettin Yesilyurt	Supervise On	9/11/2024 7:08:33 PM
SubDirectory	VN091024	HP Acquire Method	HP Processing Method 82N091024W.M
STD. NAME	STD REF.#		
Tune/Reschk	VP130211,VP130225		
Initial Calibration Stds	VP130230,VP130231,VP130232,VP130233,VP130234,VP130235		
CCC	VP130212,VP130213,VP130226		
Internal Standard/PEM			
ICV/I.BLK	VP130236		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

38	PB163204ZHE#08	PB163204ZHE#08	VN083778.D	11 Sep 2024 07:38		JC\MD	Ok,M
39	PB163204ZHE#09	PB163204ZHE#09	VN083779.D	11 Sep 2024 08:02		JC\MD	Ok,M
40	PB163204ZHE#10	PB163204ZHE#10	VN083780.D	11 Sep 2024 08:26		JC\MD	Ok,M
41	PB163204ZHE#11	PB163204ZHE#11	VN083781.D	11 Sep 2024 08:50		JC\MD	Ok,M
42	PB163204ZHE#12	PB163204ZHE#12	VN083782.D	11 Sep 2024 09:14		JC\MD	Ok
43	P3872-01	WC-S-07-202409	VN083783.D	11 Sep 2024 09:38	vial A pH#5.0	JC\MD	Ok
44	P3892-02	ARS20-005	VN083784.D	11 Sep 2024 10:02	vial A pH#5.0	JC\MD	Ok,M
45	P3906-02	SU-701-COMP-01	VN083785.D	11 Sep 2024 10:26	vial A pH#5.0	JC\MD	Ok
46	P3906-08	SU-701-COMP-02	VN083786.D	11 Sep 2024 10:50	vial A pH#5.0	JC\MD	Ok
47	P3906-14	SU-701-COMP-03	VN083787.D	11 Sep 2024 11:14	vial A pH#5.0	JC\MD	Ok
48	P3891-02	AUD-SBG	VN083788.D	11 Sep 2024 11:38	vial A pH#5.0 Surrogate Fail	JC\MD	ReRun
49	VSTDCCC050	VSTDCCC050EC	VN083789.D	11 Sep 2024 12:02		JC\MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_N

Daily Analysis Runlog For Sequence/QC Batch ID # VN091124

Review By	John Carlone	Review On	9/12/2024 9:49:59 AM
Supervise By	Maresh Dadoda	Supervise On	9/12/2024 4:33:31 PM
SubDirectory	VN091124	HP Acquire Method	HP Processing Method 82N091024W.M

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP130256
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP130257,VP130258

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VN083790.D	11 Sep 2024 12:25		JC\MD	Ok
2	VSTDCCC050	VSTDCCC050	VN083791.D	11 Sep 2024 12:59	V13516	JC\MD	Ok,M
3	VN0911MBL01	VN0911MBL01	VN083792.D	11 Sep 2024 13:33		JC\MD	Ok
4	VN0911WBL01	VN0911WBL01	VN083793.D	11 Sep 2024 13:57		JC\MD	Ok
5	P3839-02	830	VN083794.D	11 Sep 2024 14:21	vial A pH<2 Surrogate Fail	JC\MD	ReRun
6	P3839-01	828	VN083795.D	11 Sep 2024 14:45	vial A pH<2	JC\MD	Ok
7	IBLK	IBLK	VN083796.D	11 Sep 2024 15:09		JC\MD	Ok
8	P3845-01DL	PT-VOA-WPDL	VN083797.D	11 Sep 2024 15:33		JC\MD	Ok
9	IBLK	IBLK	VN083798.D	11 Sep 2024 16:21		JC\MD	Ok
10	VN0911WBS01	VN0911WBS01	VN083799.D	11 Sep 2024 16:45		JC\MD	Ok,M
11	VN0911WBSD01	VN0911WBSD01	VN083800.D	11 Sep 2024 17:09		JC\MD	Ok,M
12	P3839-02	830	VN083801.D	11 Sep 2024 17:33	vial B pH<2	JC\MD	Ok,M
13	P3891-02	AUD-SBG	VN083802.D	11 Sep 2024 17:57	vial B pH#5.0	JC\MD	Ok,M
14	P3913-04	9424-A-C	VN083803.D	11 Sep 2024 18:21	Need straight run	JC\MD	Not Ok
15	P3913-06	9424-D-F	VN083804.D	11 Sep 2024 18:45	Need straight run	JC\MD	Not Ok
16	P3913-08	9424-D-F-82924-D	VN083805.D	11 Sep 2024 19:09	Need straight run	JC\MD	Not Ok
17	P3913-10	82924-A-C	VN083806.D	11 Sep 2024 19:33	Need straight run	JC\MD	Not Ok
18	PB163246TB	PB163246TB	VN083807.D	11 Sep 2024 19:57		JC\MD	Ok,M

Instrument ID: MSVOA_N

Daily Analysis Runlog For Sequence/QC Batch ID # VN091124

Review By	John Carlone	Review On	9/12/2024 9:49:59 AM
Supervise By	Mahesh Dadoda	Supervise On	9/12/2024 4:33:31 PM
SubDirectory	VN091124	HP Acquire Method	HP Processing Method 82N091024W.M
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	VP130256		
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP130257,VP130258		

19	P3911-02	SP-1	VN083808.D	11 Sep 2024 20:21	vial A pH#5.0	JC\MD	Ok,M
20	P3911-04	SP-2	VN083809.D	11 Sep 2024 20:45	vial A pH#5.0	JC\MD	Ok,M
21	P3911-06	SP-3	VN083810.D	11 Sep 2024 21:09	vial A pH#5.0	JC\MD	Ok,M
22	P3905-06	COMP-1	VN083811.D	11 Sep 2024 21:33	vial A pH#5.0 Surrogate Fail	JC\MD	ReRun
23	P3912-01	282-COMPOSITE-1	VN083812.D	11 Sep 2024 21:57	vial A pH<2	JC\MD	Ok
24	P3913-05	9424-A-C	VN083813.D	11 Sep 2024 22:21	vial A pH#5.0	JC\MD	Ok
25	P3913-07	9424-D-F	VN083814.D	11 Sep 2024 22:45	vial A pH#5.0	JC\MD	Ok
26	P3913-09	9424-D-F-82924-D	VN083815.D	11 Sep 2024 23:09	vial A pH#5.0	JC\MD	Ok
27	P3913-11	82924-A-C	VN083816.D	11 Sep 2024 23:33	vial A pH#5.0	JC\MD	Ok
28	VSTDCCC050	VSTDCCC050EC	VN083817.D	11 Sep 2024 23:56		JC\MD	Ok,M

M : Manual Integration