

Surrogate Summary

SDG No.: **P3845**

Client: **Chemtech Consulting Group**

Analytical Method: **SW8260-Low**

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
P3845-01	PT-VOA-WP	1,2-Dichloroethane-d4	50	48.1	96	74	125
		Dibromofluoromethane	50	45.7	91	75	124
		Toluene-d8	50	48.6	97	86	113
		4-Bromofluorobenzene	50	44.6	89	77	121
P3845-01DL	PT-VOA-WPDL	1,2-Dichloroethane-d4	50	52.3	105	74	125
		Dibromofluoromethane	50	49.0	98	75	124
		Toluene-d8	50	51.0	102	86	113
		4-Bromofluorobenzene	50	49.1	98	77	121
VN0910WBL01	VN0910WBL01	1,2-Dichloroethane-d4	50	52.5	105	74	125
		Dibromofluoromethane	50	48.0	96	75	124
		Toluene-d8	50	48.1	96	86	113
		4-Bromofluorobenzene	50	53.5	107	77	121
VN0910WBS01	VN0910WBS01	1,2-Dichloroethane-d4	50	50.0	100	74	125
		Dibromofluoromethane	50	48.3	97	75	124
		Toluene-d8	50	50.8	102	86	113
		4-Bromofluorobenzene	50	47.3	95	77	121
VN0910WBSD0	VN0910WBSD01	1,2-Dichloroethane-d4	50	46.9	94	74	125
		Dibromofluoromethane	50	45.8	92	75	124
		Toluene-d8	50	48.4	97	86	113
		4-Bromofluorobenzene	50	45.1	90	77	121
VN0911WBL01	VN0911WBL01	1,2-Dichloroethane-d4	50	55.7	111	74	125
		Dibromofluoromethane	50	49.4	99	75	124
		Toluene-d8	50	51.1	102	86	113
		4-Bromofluorobenzene	50	57.0	114	77	121
VN0911WBS01	VN0911WBS01	1,2-Dichloroethane-d4	50	50.6	101	74	125
		Dibromofluoromethane	50	48.7	97	75	124
		Toluene-d8	50	51.5	103	86	113
		4-Bromofluorobenzene	50	47.0	94	77	121