

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: CHEM02
 Lab Code: CHEM Case No.: P3845 SAS No.: P3845 SDG NO.: P3845
 Lab File ID: VN083743.D Date Analyzed: 09/10/2024
 Instrument ID: MSVOA_N Time Analyzed: 09:32
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	127433	13.788				
UPPER LIMIT	254866	14.288				
LOWER LIMIT	63716.5	13.288				
EPA SAMPLE NO.						
PT-VOA-WP	109081	13.79				
VN0910WBL01	126892	13.79				
VN0910WBS01	118496	13.79				
VN0910WBSD01	112657	13.79				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.