

metals
- 5a -
MATRIX SPIKE SUMMARY

client: Chemtech Consulting Group **level:** low **sdg no.:** P3846
contract: CHEM02 **lab code:** CHEM **case no.:** P3846 **sas no.:** P3846
matrix: Water **sample id:** P4289-02 **client id:** EFF-Waste WaterMS
Percent Solids for Sample: NA **Spiked ID:** P4289-02MS **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Mercury	ug/L	75 - 125	3.68		0.034	J	4.0	91		CV

metals
- 5a -
MATRIX SPIKE DUPLICATE SUMMARY

client:	<u>Chemtech Consulting Group</u>	level:	<u>low</u>	sdg no.:	<u>P3846</u>		
contract:	<u>CHEM02</u>	lab code:	<u>CHEM</u>	case no.:	<u>P3846</u>	sas no.:	<u>P3846</u>
matrix:	<u>Water</u>	sample id:	<u>P4289-02</u>	client id:	<u>EFF-Waste WaterMSD</u>		
Percent Solids for Sample:	<u>NA</u>	Spiked ID:	<u>P4289-02MSD</u>	Percent Solids for Spike Sample:	<u>NA</u>		

Analyte	Units	Acceptance Limit %R	MSD Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Mercury	ug/L	75 - 125	3.80		0.034	J	4.0	94		CV