

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Chemtech Consulting Group

SDG No.: P3846

Contract: CHEM02

Lab Code: CHEM

Case No.: P3846

SAS No.: P3846

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB163988BS Mercury	ug/L	4.0	3.70		92	85 - 115	CV

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Chemtech Consulting Group **SDG No.:** P3846
Contract: CHEM02 **Lab Code:** CHEM **Case No.:** P3846 **SAS No.:** P3846

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB164232BS							
Aluminum	ug/L	2000	1980		99	85 - 115	P
Antimony	ug/L	800	813		102	85 - 115	P
Arsenic	ug/L	800	830		104	85 - 115	P
Barium	ug/L	200	197		98	85 - 115	P
Beryllium	ug/L	200	204		102	85 - 115	P
Boron	ug/L	300	293		98	85 - 115	P
Cadmium	ug/L	200	204		102	85 - 115	P
Calcium	ug/L	1000	1040	J	104	85 - 115	P
Chromium	ug/L	400	425		106	85 - 115	P
Cobalt	ug/L	200	204		102	85 - 115	P
Copper	ug/L	300	317		106	85 - 115	P
Iron	ug/L	3000	3060		102	85 - 115	P
Lead	ug/L	1000	1030		103	85 - 115	P
Magnesium	ug/L	2000	1970	J	98	85 - 115	P
Manganese	ug/L	200	202		101	85 - 115	P
Molybdenum	ug/L	400	415		104	85 - 115	P
Nickel	ug/L	500	514		103	85 - 115	P
Potassium	ug/L	10000	10100		101	85 - 115	P
Selenium	ug/L	2000	2020		101	85 - 115	P
Silicon	ug/L	930	837		90	85 - 115	P
Silver	ug/L	75.0	78.0		104	85 - 115	P
Sodium	ug/L	3000	2930		98	85 - 115	P
Strontium	ug/L	200	205		102	85 - 115	P
Thallium	ug/L	2000	2100		105	85 - 115	P
Tin	ug/L	700	699		100	85 - 115	P
Titanium	ug/L	200	195		98	85 - 115	P
Vanadium	ug/L	300	298		99	85 - 115	P
Zinc	ug/L	200	215		108	85 - 115	P