

Metals
- 3b -
PREPARATION BLANK SUMMARY

Client: Chemtech Consulting Group

SDG No.: P3846

Instrument: P7

Sample ID	Analyte	Result (ug/L)	Acceptance Limit	Conc Qual	CRQL ug/L	M	Analysis Date	Analysis Time	Run
PB164204BL	WATER			Batch Number:	PB164204		Prep Date:	10/17/2024	
	Aluminum	20.0	<20.0	U	20.0	P	10/17/2024	19:22	LB132993
	Antimony	0.32	<2.00	J	2.00	P	10/17/2024	19:22	LB132993
	Arsenic	1.00	<1.00	U	1.00	P	10/17/2024	19:22	LB132993
	Barium	10.0	<10.0	U	10.0	P	10/17/2024	19:22	LB132993
	Beryllium	0.26	<1.00	J	1.00	P	10/17/2024	19:22	LB132993
	Cadmium	1.00	<1.00	U	1.00	P	10/17/2024	19:22	LB132993
	Calcium	500	<500	U	500	P	10/17/2024	19:22	LB132993
	Chromium	2.00	<2.00	U	2.00	P	10/17/2024	19:22	LB132993
	Cobalt	1.00	<1.00	U	1.00	P	10/17/2024	19:22	LB132993
	Copper	2.00	<2.00	U	2.00	P	10/17/2024	19:22	LB132993
	Iron	50.0	<50.0	U	50.0	P	10/17/2024	19:22	LB132993
	Lead	1.00	<1.00	U	1.00	P	10/17/2024	19:22	LB132993
	Magnesium	500	<500	U	500	P	10/17/2024	19:22	LB132993
	Manganese	1.00	<1.00	U	1.00	P	10/17/2024	19:22	LB132993
	Nickel	1.00	<1.00	U	1.00	P	10/17/2024	19:22	LB132993
	Potassium	500	<500	U	500	P	10/17/2024	19:22	LB132993
	Tin	5.00	<5.00	U	5.00	P	10/17/2024	19:22	LB132993
	Molybdenum	5.00	<5.00	U	5.00	P	10/17/2024	19:22	LB132993
	Selenium	5.00	<5.00	U	5.00	P	10/17/2024	19:22	LB132993
	Silver	1.00	<1.00	U	1.00	P	10/17/2024	19:22	LB132993
	Sodium	500	<500	U	500	P	10/17/2024	19:22	LB132993
	Thallium	1.00	<1.00	U	1.00	P	10/17/2024	19:22	LB132993
	Vanadium	5.00	<5.00	U	5.00	P	10/17/2024	19:22	LB132993
	Zinc	5.00	<5.00	U	5.00	P	10/17/2024	19:22	LB132993
	Titanium	0.29	<5.00	J	5.00	P	10/17/2024	19:22	LB132993