

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Chemtech Consulting Group **SDG No.:** P3846
Contract: CHEM02 **Lab Code:** CHEM **Case No.:** P3846 **SAS No.:** P3846

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB164204BS							
Aluminum	ug/L	10000	10200		102	85 - 115	P
Antimony	ug/L	500	517		103	85 - 115	P
Arsenic	ug/L	500	483		97	85 - 115	P
Barium	ug/L	2500	2540		102	85 - 115	P
Beryllium	ug/L	500	496		99	85 - 115	P
Cadmium	ug/L	500	511		102	85 - 115	P
Calcium	ug/L	50000	53700		107	85 - 115	P
Chromium	ug/L	500	500		100	85 - 115	P
Cobalt	ug/L	500	492		98	85 - 115	P
Copper	ug/L	5000	4970		99	85 - 115	P
Iron	ug/L	25000	28200		113	85 - 115	P
Lead	ug/L	2500	2520		101	85 - 115	P
Magnesium	ug/L	50000	48100		96	85 - 115	P
Manganese	ug/L	5000	4890		98	85 - 115	P
Molybdenum	ug/L	5000	5550		111	85 - 115	P
Nickel	ug/L	500	494		99	85 - 115	P
Potassium	ug/L	25000	25500		102	85 - 115	P
Selenium	ug/L	500	504		101	85 - 115	P
Silver	ug/L	500	516		103	85 - 115	P
Sodium	ug/L	50000	51300		103	85 - 115	P
Thallium	ug/L	500	508		102	85 - 115	P
Tin	ug/L	500	515		103	85 - 115	P
Titanium	ug/L	5000	5610		112	85 - 115	P
Vanadium	ug/L	500	486		97	85 - 115	P
Zinc	ug/L	5000	4910		98	85 - 115	P