

FORM 8B

ICP-MS INTERNAL STANDARD RELATIVE INTENSITY SUMMARY

Lab Name: Chemtech Consulting Group

Contract: CHEM02

Lab Code: CHEM Case no.: P3846

Sas No.: P3846 SDG No.: P3846

Instrument ID: P7

Start Date : 10/17/2024

Run Number: LB132993

End Date : 10/17/2024

Lab SampleID	Client SampleID	Time	Internal Standard %RI For: Non-Collision Cell									
			Element 165Ho	Q	Element 209Bi	Q	Element	Q	Element	Q	Element	Q
S0	S0	1619	100		100							
S2	S2	1623	101		100							
S3	S3	1626	102		101							
S4	S4	1629	102		101							
S5	S5	1633	102		101							
S6	S6	1636	103		100							
S7	S7	1641	108		104							
S8	S8	1644	100		90							
ICV01	ICV01	1703	107		107							
LLICV	LLICV	1709	107		108							
ICB01	ICB01	1717	104		104							
ICSA01	ICSA01	1720	106		100							
ICSAB01	ICSAB01	1723	109		103							
CCV01	CCV01	1804	105		96							
CCB01	CCB01	1807	108		109							
CRI	CRI	1811	107		107							
ZZZZZZ	ZZZZZZ	1825										
ZZZZZZ	ZZZZZZ	1828										
ZZZZZZ	ZZZZZZ	1832										
P3846-02	PT-TM1-WP	1835	105		106							
P3846-04	PT-SNTI-WP	1838	106		104							
P3846-06	PT-MIN2-WP	1841	153	*	145	*						
ZZZZZZ	ZZZZZZ	1845										
ZZZZZZ	ZZZZZZ	1848										
ZZZZZZ	ZZZZZZ	1851										
CCV02	CCV02	1855	103		94							
CCB02	CCB02	1858	106		105							
P3846-02	PT-TM1-WP	1901	107		108							
P3846-04	PT-SNTI-WP	1904	106		106							
P3846-06	PT-MIN2-WP	1908	120		116							
ZZZZZZ	ZZZZZZ	1916										
ZZZZZZ	ZZZZZZ	1920										
PB164204BL	PB164204BL	1922	107		106							
PB164204BS	PB164204BS	1926	107		103							

Internal Standard %RI Limit: 70 -130

FORM 8B

ICP-MS INTERNAL STANDARD RELATIVE INTENSITY SUMMARY

Lab Name: Chemtech Consulting Group

Contract: CHEM02

Lab Code: CHEM Case no.: P3846

Sas No.: P3846 SDG No.: P3846

Instrument ID: P7

Start Date : 10/17/2024

Run Number: LB132993

End Date : 10/17/2024

Lab SampleID	Client SampleID	Time	Internal Standard %RI For: Non-Collision Cell									
			Element 165Ho	Q	Element 209Bi	Q	Element	Q	Element	Q	Element	Q
CCV03	CCV03	1929	105		96							
CCB03	CCB03	1935	106		106							

FORM 8B

ICP-MS INTERNAL STANDARD RELATIVE INTENSITY SUMMARY

Lab Name: Chemtech Consulting Group

Contract: CHEM02

Lab Code: CHEM Case no.: P3846

Sas No.: P3846 SDG No.: P3846

Instrument ID: P7

Start Date : 10/17/2024

Run Number: LB132993

End Date : 10/17/2024

Lab SampleID	Client SampleID	Time	Internal Standard %RI For: Collision Cell									
			Element 209Bi	Q	Element	Q	Element	Q	Element	Q	Element	Q
S0	S0	1619	100									
S2	S2	1623	104									
S3	S3	1626	105									
S4	S4	1629	105									
S5	S5	1633	104									
S6	S6	1636	106									
S7	S7	1641	99									
S8	S8	1644	84									
ICV01	ICV01	1703	106									
LLICV	LLICV	1709	110									
ICB01	ICB01	1717	110									
ICSA01	ICSA01	1720	105									
ICSAB01	ICSAB01	1723	106									
CCV01	CCV01	1804	88									
CCB01	CCB01	1807	113									
CRI	CRI	1811	113									
ZZZZZZ	ZZZZZZ	1825										
ZZZZZZ	ZZZZZZ	1828										
ZZZZZZ	ZZZZZZ	1832										
P3846-02	PT-TM1-WP	1835	110									
P3846-04	PT-SNTI-WP	1838	111									
P3846-06	PT-MIN2-WP	1841	141	*								
ZZZZZZ	ZZZZZZ	1845										
ZZZZZZ	ZZZZZZ	1848										
ZZZZZZ	ZZZZZZ	1851										
CCV02	CCV02	1855	88									
CCB02	CCB02	1858	105									
P3846-02	PT-TM1-WP	1901	113									
P3846-04	PT-SNTI-WP	1904	111									
P3846-06	PT-MIN2-WP	1908	119									
ZZZZZZ	ZZZZZZ	1916										
ZZZZZZ	ZZZZZZ	1920										
PB164204BL	PB164204BL	1922	110									
PB164204BS	PB164204BS	1926	107									

Internal Standard %RI Limit: 70 -130

FORM 8B

ICP-MS INTERNAL STANDARD RELATIVE INTENSITY SUMMARY

Lab Name: Chemtech Consulting Group

Contract: CHEM02

Lab Code: CHEM Case no.: P3846

Sas No.: P3846 SDG No.: P3846

Instrument ID: P7

Start Date : 10/17/2024

Run Number: LB132993

End Date : 10/17/2024

Lab SampleID	Client SampleID	Time	Internal Standard %RI For: Collision Cell									
			Element 209Bi	Q	Element	Q	Element	Q	Element	Q	Element	Q
CCV03	CCV03	1929	89									
CCB03	CCB03	1935	111									