

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Chemtech Consulting Group **SDG No.:** P3847
Contract: CHEM02 **Lab Code:** CHEM **Case No.:** P3847 **SAS No.:** P3847

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB164203BS							
Aluminum	ug/L	10000	10100		101	80 - 120	P
Antimony	ug/L	500	552		110	80 - 120	P
Arsenic	ug/L	500	480		96	80 - 120	P
Barium	ug/L	2500	2690		108	80 - 120	P
Beryllium	ug/L	500	527		105	80 - 120	P
Cadmium	ug/L	500	548		110	80 - 120	P
Calcium	ug/L	50000	56000		112	80 - 120	P
Chromium	ug/L	500	489		98	80 - 120	P
Cobalt	ug/L	500	486		97	80 - 120	P
Copper	ug/L	5000	4900		98	80 - 120	P
Iron	ug/L	25000	27600		110	80 - 120	P
Lead	ug/L	2500	2680		107	80 - 120	P
Magnesium	ug/L	50000	47600		95	80 - 120	P
Manganese	ug/L	5000	4800		96	80 - 120	P
Nickel	ug/L	500	485		97	80 - 120	P
Potassium	ug/L	25000	25300		101	80 - 120	P
Tin	ug/L	500	547		109	80 - 120	P
Molybdenum	ug/L	5000	5890		118	80 - 120	P
Selenium	ug/L	500	543		109	80 - 120	P
Silver	ug/L	500	557		111	80 - 120	P
Sodium	ug/L	50000	50800		102	80 - 120	P
Thallium	ug/L	500	536		107	80 - 120	P
Vanadium	ug/L	500	476		95	80 - 120	P
Zinc	ug/L	5000	4840		97	80 - 120	P
Strontium	ug/L	12500	13400		107	80 - 120	P
Titanium	ug/L	5000	5900		118	80 - 120	P