

FORM 8B

ICP-MS INTERNAL STANDARD RELATIVE INTENSITY SUMMARY

Lab Name: Chemtech Consulting Group

Contract: CHEM02

Lab Code: CHEM Case no.: P3847

Sas No.: P3847 SDG No.: P3847

Instrument ID: P7

Start Date : 10/17/2024

Run Number: LB132993

End Date : 10/17/2024

Lab SampleID	Client SampleID	Time	Internal Standard %RI For: Non-Collision Cell									
			Element 165Ho	Q	Element 209Bi	Q	Element	Q	Element	Q	Element	Q
S0	S0	1619	100		100							
S2	S2	1623	101		100							
S3	S3	1626	102		101							
S4	S4	1629	102		101							
S5	S5	1633	102		101							
S6	S6	1636	103		100							
S7	S7	1641	108		104							
S8	S8	1644	100		90							
ICV01	ICV01	1703	107		107							
LLICV	LLICV	1709	107		108							
ICB01	ICB01	1717	104		104							
ICSA01	ICSA01	1720	106		100							
ICSAB01	ICSAB01	1723	109		103							
CCV01	CCV01	1804	105		96							
CCB01	CCB01	1807	108		109							
CRI	CRI	1811	107		107							
P3847-02	PT-TM1-WP	1825	108		108							
P3847-04	PT-SNTI-WP	1828	111		109							
P3847-06	PT-MIN2-WP	1832	150	*	143	*						
ZZZZZZ	ZZZZZZ	1835										
ZZZZZZ	ZZZZZZ	1838										
ZZZZZZ	ZZZZZZ	1841										
P3847-02	PT-TM1-WP	1845	104		103							
P3847-04	PT-SNTI-WP	1848	105		104							
P3847-06	PT-MIN2-WP	1851	110		109							
CCV02	CCV02	1855	103		94							
CCB02	CCB02	1858	106		105							
ZZZZZZ	ZZZZZZ	1901										
ZZZZZZ	ZZZZZZ	1904										
ZZZZZZ	ZZZZZZ	1908										
PB164203BL	PB164203BL	1916	106		107							
PB164203BS	PB164203BS	1920	101		96							
ZZZZZZ	ZZZZZZ	1922										
ZZZZZZ	ZZZZZZ	1926										

Internal Standard %RI Limit: 70 -130

FORM 8B

ICP-MS INTERNAL STANDARD RELATIVE INTENSITY SUMMARY

Lab Name: Chemtech Consulting Group

Contract: CHEM02

Lab Code: CHEM Case no.: P3847

Sas No.: P3847 SDG No.: P3847

Instrument ID: P7

Start Date : 10/17/2024

Run Number: LB132993

End Date : 10/17/2024

Lab SampleID	Client SampleID	Time	Internal Standard %RI For: Non-Collision Cell									
			Element 165Ho	Q	Element 209Bi	Q	Element	Q	Element	Q	Element	Q
CCV03	CCV03	1929	105		96							
CCB03	CCB03	1935	106		106							

FORM 8B

ICP-MS INTERNAL STANDARD RELATIVE INTENSITY SUMMARY

Lab Name: Chemtech Consulting Group

Contract: CHEM02

Lab Code: CHEM Case no.: P3847

Sas No.: P3847 SDG No.: P3847

Instrument ID: P7

Start Date : 10/17/2024

Run Number: LB132993

End Date : 10/17/2024

Lab SampleID	Client SampleID	Time	Internal Standard %RI For: Collision Cell									
			Element 209Bi	Q	Element	Q	Element	Q	Element	Q	Element	Q
S0	S0	1619	100									
S2	S2	1623	104									
S3	S3	1626	105									
S4	S4	1629	105									
S5	S5	1633	104									
S6	S6	1636	106									
S7	S7	1641	99									
S8	S8	1644	84									
ICV01	ICV01	1703	106									
LLICV	LLICV	1709	110									
ICB01	ICB01	1717	110									
ICSA01	ICSA01	1720	105									
ICSAB01	ICSAB01	1723	106									
CCV01	CCV01	1804	88									
CCB01	CCB01	1807	113									
CRI	CRI	1811	113									
P3847-02	PT-TM1-WP	1825	111									
P3847-04	PT-SNTI-WP	1828	114									
P3847-06	PT-MIN2-WP	1832	145	*								
ZZZZZZ	ZZZZZZ	1835										
ZZZZZZ	ZZZZZZ	1838										
ZZZZZZ	ZZZZZZ	1841										
P3847-02	PT-TM1-WP	1845	110									
P3847-04	PT-SNTI-WP	1848	111									
P3847-06	PT-MIN2-WP	1851	112									
CCV02	CCV02	1855	88									
CCB02	CCB02	1858	105									
ZZZZZZ	ZZZZZZ	1901										
ZZZZZZ	ZZZZZZ	1904										
ZZZZZZ	ZZZZZZ	1908										
PB164203BL	PB164203BL	1916	111									
PB164203BS	PB164203BS	1920	106									
ZZZZZZ	ZZZZZZ	1922										
ZZZZZZ	ZZZZZZ	1926										

Internal Standard %RI Limit: 70 -130

FORM 8B

ICP-MS INTERNAL STANDARD RELATIVE INTENSITY SUMMARY

Lab Name: Chemtech Consulting Group

Contract: CHEM02

Lab Code: CHEM Case no.: P3847

Sas No.: P3847 SDG No.: P3847

Instrument ID: P7

Start Date : 10/17/2024

Run Number: LB132993

End Date : 10/17/2024

Lab SampleID	Client SampleID	Time	Internal Standard %RI For: Collision Cell									
			Element 209Bi	Q	Element	Q	Element	Q	Element	Q	Element	Q
CCV03	CCV03	1929	89									
CCB03	CCB03	1935	111									