

Analytical Summary Report

Analysis Method: SM2510 B
Parameter: Conductance
Run Number: LB132420

ANALYST: Niha
SUPERVISOR REVIEW BY: Iwona
Meter ID: WC CONDUCTANCE METER-1
Thermometer ID: WC PH METER-1

Reagent/Standard	Lot/Log #
conductivity standard	WP107551

Seq	Lab ID	True Value (mg/L)	Dilution	Temperature °C	Reading (umho/cm)	Result (umho/cm)	Anal Date	Anal Time
1	LB132420BL		1	23.0	0.92	0.96	09/13/2024	12:30
2	LB132420BS	1413	1	23.0	1372.00	1426.49	09/13/2024	12:35
3	P3848-07		1	23.5	753.00	775.21	09/13/2024	12:40

Note : Conductivity result is corrected to 25°C

Result = (Reading)
1 + 0.0191 (Temperature - 25)

LB 132420

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CONDUCTANCE-091320

WorkList ID : 183481

Department : Wet-Chemistry

Date : 09-13-2024 11:16:13

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P3848-07	PT-COND-WP	Water	Conductance	Cool 4 deg C	CHEM02	QA Of	09/03/2024	SM2510 B

Date/Time 09.13.2024 , 11:30
Raw Sample Received by: NF (LWC)
Raw Sample Relinquished by: SJ (GAO)

Date/Time N/A
Raw Sample Received by: "
Raw Sample Relinquished by: "