

Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB132548

Slope : 98.9

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3107
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER, COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	09/23/2024	11:10
2	CAL2	1	Water	NA	NA	20.2	7.00	09/23/2024	11:11
3	CAL3	1	Water	NA	NA	20.3	10.03	09/23/2024	11:12
4	ICV	1	Water	NA	NA	20.3	7.00	09/23/2024	11:15
5	CCV1	1	Water	NA	NA	20.2	2.01	09/23/2024	11:20
6	P3848-20	1	Water	NA	NA	20.2	5.37	09/23/2024	11:25
7	P4140-02	1	Water	NA	NA	24.6	7.40	09/23/2024	11:30
8	P4140-02DUP	1	Water	NA	NA	24.7	7.41	09/23/2024	11:31
9	CCV2	1	Water	NA	NA	20.3	12.02	09/23/2024	11:35

WORKLIST(Hardcopy Internal Chain)

132548

WorkList Name : PH P4144 WorkList ID : 183727 Department : Wet-Chemistry Date : 09-23-2024 10:56:28

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P3848-20	PT-PH-WP	Water	pH	Cool 4 deg C	CHEM02	QA Of	09/03/2024	9040C
P4140-02	306-307	Water	pH	Cool 4 deg C	PSEG03	J31	09/20/2024	9040C

Date/Time 09/23/24 11:00
Raw Sample Received by: JG
Raw Sample Relinquished by: RM SM

Date/Time 09/23/24 13:00
Raw Sample Received by: RM SM
Raw Sample Relinquished by: RM WLS