

Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: Niha

Parameter: Turbidity

SUPERVISOR REVIEW BY: Iwona

Run Number: LB132624

Reagent/Standard	Lot/Log #
Turbidity Calibration std, 0NTU	WP109925
Turbidity Calibration std, 20NTU	WP109927
Turbidity Calibration std, 1NTU	WP109929
Turbidity Calibration std, 5NTU	WP109928
Turbidity Calibration - CCV std, 10 NTU	WP109930
10 NTU Standard 500 ml	W3116

Intercept: -0.0156

Slope: 0.9363

Regression: 0.9999

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (mg/l)	%D	Anal Date	Anal Time
1	CAL1	0	1	0.148	0.18		09/26/2024	10:24
2	CAL2	1	1	1.020	1.11	10.6	09/26/2024	10:28
3	CAL3	5	1	4.510	4.83	-3.3	09/26/2024	10:32
4	CAL4	10	1	9.390	10.05	0.4	09/26/2024	10:36
5	CAL5	20	1	18.400	19.67	-1.7	09/26/2024	10:40
6	CAL6	40	1	37.600	40.18	0.4	09/26/2024	10:44



Analytical Summary Report

Reviewed By: Iwona
On: 9/30/2024 11:53:08 AM
Inst Id: WC
TURBIDIMETER-1

Analysis Method: SM2130 B

ANALYST: Niha

Parameter: Turbidity

SUPERVISOR REVIEW BY: Iwona

Run Number: LB132624

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (NTU)	AnalDate	Anal Time
1	ICV	10	1	8.910	9.533	09/26/2024	10:48
2	ICB		1	0.146	0.173	09/26/2024	10:52
3	CCV1	10	1	8.820	9.437	09/26/2024	10:56
4	CCB1		1	0.141	0.167	09/26/2024	11:00
5	LB132624BL		1	0.138	0.164	09/26/2024	11:04
6	P3848-33		1	24.100	25.756	09/26/2024	11:08
7	P4200-01		1	0.810	0.882	09/26/2024	11:24
8	P4200-01DUP		1	0.800	0.871	09/26/2024	11:28
9	P4200-02		1	0.600	0.657	09/26/2024	11:32
10	P4200-03		1	0.500	0.551	09/26/2024	11:36
11	P4200-04		1	1.720	1.854	09/26/2024	11:40
12	P4200-05		1	0.595	0.652	09/26/2024	11:44
13	P4200-06		1	0.568	0.623	09/26/2024	11:48
14	P4200-07		1	0.690	0.754	09/26/2024	11:52
15	CCV2	10	1	9.200	9.843	09/26/2024	11:56
16	CCB2		1	0.142	0.168	09/26/2024	12:00

LB132624

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TURBIDITY-09262024 WorkList ID : 183852 Department : Wet-Chemistry Date : 09-26-2024 09:03:28

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P4200-01	B2-WARD-12-RM3016-SINK	Water	Turbidity	Cool 4 deg C	CLOU03	J41	09/25/2024	SM2130 B
P4200-02	B2-WARD-12-RM3010-SINK	Water	Turbidity	Cool 4 deg C	CLOU03	J41	09/25/2024	SM2130 B
P4200-03	B2-WARD-12-RM3011-RIGHT-SINK	Water	Turbidity	Cool 4 deg C	CLOU03	J41	09/25/2024	SM2130 B
P4200-04	B2-WARD-12-LEFT-DRINKING	Water	Turbidity	Cool 4 deg C	CLOU03	J41	09/25/2024	SM2130 B
P4200-05	B2-WARD-12-RM3019-EYEWA	Water	Turbidity	Cool 4 deg C	CLOU03	J41	09/25/2024	SM2130 B
P4200-06	B2-WARD-12-RM3040-SINK	Water	Turbidity	Cool 4 deg C	CLOU03	J41	09/25/2024	SM2130 B
P4200-07	B2-WARD-12-RM3035-SINK	Water	Turbidity	Cool 4 deg C	CLOU03	J41	09/25/2024	SM2130 B
P3848-33	PT-TURB-WP	Water	Turbidity	Cool 4 deg C	CHEM02	QA Of	09/03/2024	SM2130 B

Date/Time 09.26.2024, 09:10
Raw Sample Received by: NF(wc)
Raw Sample Relinquished by: [Signature]

Date/Time 09.26.2024, 12:00
Raw Sample Received by: [Signature]
Raw Sample Relinquished by: NF(wc)