

TOTAL VOLATILE SOLIDS 160.4

TEMP1 IN: 103 °C 10/15/2024 16:00 **TEMP1 OUT:** 103 °C 10/16/2024 07:30
TEMP2 IN: 104 °C 10/16/2024 08:00 **TEMP2 OUT:** 103 °C 10/16/2024 09:30
TEMP3 IN: 540 °C 10/16/2024 10:00 **TEMP3 OUT:** 560 °C 10/16/2024 11:30
TEMP4 IN: 540 °C 10/16/2024 12:00 **TEMP4 OUT:** 540 °C 10/16/2024 13:30

Run Number: LB132944

SUPERVISOR: Iwona

ANALYST: jignesh

BalanceID: WC SC-6

OvenID: WC OVEN-1

Dish #	Lab ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Vol (ml)	1st Dish + SampleWt Drying @103-@105°C (g)	Final Dish + SampleWt Drying @103-@105°C (g)	Dish + Samplewt Drying @550 (±50) °C (g)	Final Dish + Samplewt Drying @550 (±50) °C (g)	Weight Diff (g)	Result (mg/L)
1	LB132944BL	94.5841	94.5841	100	94.5841	94.5841	94.5841	94.5841	0.0000	0
2	P3848-32	94.5797	94.5797	100	94.6206	94.6206	94.5834	94.5834	0.0372	372

A = Sample Volume (ml)
B = Final Dish + Samplewt Drying @550 (±50) °C (g)
C = Final Dish + SampleWt Drying @103-@105 °C (g)
D = Weight (g)
E = Final Empty Dish Weight (g)
F = Final Dish + SampleWt Drying @103-@105 °C (g)

Weight D = C - B

Result % = $\frac{D}{A} \star 1000000$

WORKLIST(Hardcopy Internal Chain)

132944

WorkList Name : tvs p3848 water WorkList ID : 184455 Department : Wet-Chemistry Date : 10-15-2024 11:43:41

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P3848-32	PT-VSOL-WP	Water	TVS	Cool 4 deg C	CHEM02	QA Of	09/03/2024	160.4

Date/Time 10/15-2024 10:15:24 12:30
Raw Sample Received by: JO WOOD
Raw Sample Relinquished by: SJ(QAO)

Date/Time 10/15/24 16:30
Raw Sample Received by: N/A
Raw Sample Relinquished by: JO WOOD