

Instrument ID: BNA_P

Daily Analysis Runlog For Sequence/QC Batch ID # BP093024

Review By	Jagrut	Review On	10/1/2024 11:04:18 AM
Supervise By	mohammad	Supervise On	10/2/2024 9:04:47 AM
SubDirectory	BP093024	HP Acquire Method	BNA_P
		HP Processing Method	BP092424

STD. NAME	STD REF.#
Tune/Reschk	SP6633
Initial Calibration Stds	SP6585,SP6580,SP6581,SP6582,SP6583,SP6584
CCC	SP6581
Internal Standard/PEM	S12317,10ul/1000ul sample
ICV/I.BLK	SP6618
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP762	DFTPP762	BP022111.D	30 Sep 2024 13:57		RC/JU	Ok
2	SSTDCCC020	SSTD020777	BP022112.D	30 Sep 2024 14:38		RC/JU	Ok,M
3	PB163717BL	SBLK717	BP022113.D	30 Sep 2024 15:19		RC/JU	Ok,M
4	PB163716BL	SBLK716	BP022114.D	30 Sep 2024 16:00		RC/JU	Ok
5	P4022-06	DDCD6	BP022115.D	30 Sep 2024 16:41	Need 5X & 50X Dilution	RC/JU	Dilution
6	P4022-07	DDCA1	BP022116.D	30 Sep 2024 17:22	Need 5X & 50X Dilution	RC/JU	Dilution
7	P4022-08	DDCA6	BP022117.D	30 Sep 2024 18:03	Need 50X Dilution	RC/JU	Dilution
8	P4022-01	DDC37	BP022118.D	30 Sep 2024 18:44	Need 50X Dilution	RC/JU	Dilution
9	P4022-02	DDC70	BP022119.D	30 Sep 2024 19:24	Need 20X Dilution	RC/JU	Dilution
10	P4022-03MS	DDC70MS	BP022120.D	30 Sep 2024 20:05	Internal Standard Fail	RC/JU	Not Ok
11	P4022-04MSD	DDC70MSD	BP022121.D	30 Sep 2024 20:46		RC/JU	Ok,M
12	PB163716TB	SLEB650	BP022122.D	30 Sep 2024 21:26		RC/JU	Ok
13	P4022-05	DDC95	BP022123.D	30 Sep 2024 22:07	Need 30X Dilution	RC/JU	Dilution
14	PB163716BS	SLCS716	BP022124.D	30 Sep 2024 22:48	Internal Standard Fail	RC/JU	Not Ok
15	PB163717BS	SLCS717	BP022125.D	30 Sep 2024 23:29		RC/JU	Ok,M
16	SSTDCCC020EC	SSTD020778	BP022126.D	01 Oct 2024 00:50		RC/JU	Ok,M
17	PB163565BL	SBLK565	BP022127.D	01 Oct 2024 01:31	END CCC Missing	RC/JU	Not Ok

M : Manual Integration