

Instrument ID: BNA_P

Daily Analysis Runlog For Sequence/QC Batch ID # BP100224

Review By	yogesh	Review On	10/4/2024 12:33:53 AM
Supervise By	mohammad	Supervise On	10/5/2024 4:48:07 AM
SubDirectory	BP100224	HP Acquire Method	BNA_P
		HP Processing Method	BP092424

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	SP6633 SP6585,SP6580,SP6581,SP6582,SP6583,SP6584
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	SP6581 S12317,10ul/1000ul sample SP6618

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP764	DFTPP764	BP022156.D	02 Oct 2024 10:41		RC/JU	Ok
2	SSTDCCC020	SSTD020782	BP022157.D	02 Oct 2024 11:23		RC/JU	Ok,M
3	PB163456BL	SBLK456	BP022158.D	02 Oct 2024 12:04	Analyzed for contamination check	RC/JU	Not Ok
4	P4016-11	DCVW2	BP022159.D	02 Oct 2024 12:44	Need 10X & 30X dilution & Medium level extraction required	RC/JU	Dilution
5	P4016-12	DCVW3	BP022160.D	02 Oct 2024 13:25	Need 30X dilution	RC/JU	Dilution
6	P4016-01	DCVT2	BP022161.D	02 Oct 2024 14:06		RC/JU	Ok
7	P4016-02	DCVT3	BP022162.D	02 Oct 2024 14:47		RC/JU	Ok
8	P4016-03	DCVT4	BP022163.D	02 Oct 2024 15:27		RC/JU	Ok
9	P4016-04	DCVT5	BP022164.D	02 Oct 2024 16:08		RC/JU	Ok
10	P4010-06	C0AW0	BP022165.D	02 Oct 2024 16:49		RC/JU	Ok,M
11	P4016-05	DCVT6	BP022166.D	02 Oct 2024 17:29	Need 10X	RC/JU	Dilution
12	P4016-06	DCVT7	BP022167.D	02 Oct 2024 18:10		RC/JU	Ok
13	P4016-07	DCVT8	BP022168.D	02 Oct 2024 18:51		RC/JU	Ok
14	P4016-08	DCVT9	BP022169.D	02 Oct 2024 19:32		RC/JU	Ok
15	SSTDCCC020	SSTD020783	BP022170.D	02 Oct 2024 20:53		RC/JU	Ok,M
16	PB163489BL	SBLK489	BP022171.D	02 Oct 2024 21:34		RC/JU	Ok
17	P4016-09	DCVW0	BP022172.D	02 Oct 2024 22:14		RC/JU	Ok

Instrument ID: BNA_P

Daily Analysis Runlog For Sequence/QC Batch ID # BP100224

Review By	yogesh	Review On	10/4/2024 12:33:53 AM		
Supervise By	mohammad	Supervise On	10/5/2024 4:48:07 AM		
SubDirectory	BP100224	HP Acquire Method	BNA_P	HP Processing Method	BP092424
STD. NAME	STD REF.#				
Tune/Reschk	SP6633				
Initial Calibration Stds	SP6585,SP6580,SP6581,SP6582,SP6583,SP6584				
CCC	SP6581				
Internal Standard/PEM	S12317,10ul/1000ul sample				
ICV/I.BLK	SP6618				
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

18	P4016-10	DCVW1	BP022173.D	02 Oct 2024 22:55		RC/JU	Ok
19	P4010-08DL	C0AW9DL	BP022174.D	02 Oct 2024 23:36		RC/JU	Ok,M
20	P4022-01DL	DDC37DL	BP022175.D	03 Oct 2024 00:16		RC/JU	Ok,M
21	P4017-03	DDC42	BP022176.D	03 Oct 2024 00:57	Need 10X Dilution and Medium level Extraction Required	RC/JU	Dilution
22	P4017-06	DDC69	BP022177.D	03 Oct 2024 01:37	Need 30X Dilution	RC/JU	Dilution
23	P4017-07	DDC70	BP022178.D	03 Oct 2024 02:18	Need 30X dilution and also Medium level extraction Required	RC/JU	Dilution
24	P4017-08MS	DDC70MS	BP022179.D	03 Oct 2024 02:58		RC/JU	Ok,M
25	P4017-09MSD	DDC70MSD	BP022180.D	03 Oct 2024 03:39		RC/JU	Ok,M
26	P4017-11	DDC86	BP022181.D	03 Oct 2024 04:20	Need 30X dilution and also Medium level extraction Required	RC/JU	Dilution
27	P4017-17	DDCD3	BP022182.D	03 Oct 2024 05:00	EPA COMMUNICATIOIN FOR Surrogate Fail, Need 10X & 30X Dilution	RC/JU	Dilution
28	PB163489BS	SLCS489	BP022183.D	03 Oct 2024 05:41		RC/JU	Ok,M
29	SSTDCCC020	SSTD020784	BP022184.D	03 Oct 2024 07:01		RC/JU	Ok
30	PB163456BL	SBLK456	BP022185.D	03 Oct 2024 07:42	Analyzed for contamination check	RC/JU	Ok
31	P4017-14	DDC96	BP022186.D	03 Oct 2024 08:23	Need 20X Dilution and Medium level Extraction Required.	RC/JU	Dilution
32	P4017-20	DDCD6	BP022187.D	03 Oct 2024 09:04	Need 10X Dilution and Medium level Extraction Required	RC/JU	Dilution
33	P4017-02	DDC41	BP022188.D	03 Oct 2024 09:44	Need 20X Dilution. Medium level Extraction Required	RC/JU	Dilution

Instrument ID: BNA_P

Daily Analysis Runlog For Sequence/QC Batch ID # BP100224

Review By	yogesh	Review On	10/4/2024 12:33:53 AM		
Supervise By	mohammad	Supervise On	10/5/2024 4:48:07 AM		
SubDirectory	BP100224	HP Acquire Method	BNA_P	HP Processing Method	BP092424
STD. NAME	STD REF.#				
Tune/Reschk	SP6633				
Initial Calibration Stds	SP6585,SP6580,SP6581,SP6582,SP6583,SP6584				
CCC	SP6581				
Internal Standard/PEM	S12317,10ul/1000ul sample				
ICV/I.BLK	SP6618				
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

34	P4017-19	DDCD5	BP022189.D	03 Oct 2024 10:25	Need 10X Dilution and Medium level Extraction Required	RC/JU	Dilution
35	P4016-11DL	DCVW2DL	BP022190.D	03 Oct 2024 11:05	Medium Level Extraction required, Run 10X dilution in Medium level	RC/JU	Dilution
36	P4016-11DL2	DCVW2DL2	BP022191.D	03 Oct 2024 11:46	Internal Standard Fail	RC/JU	Not Ok
37	P4016-12DL	DCVW3DL	BP022192.D	03 Oct 2024 12:27	Medium Level Extraction required, Run 10X dilution in Medium level	RC/JU	Dilution
38	P4016-05DL	DCVT6DL	BP022193.D	03 Oct 2024 13:07		RC/JU	Ok
39	P4016-05DL2	DCVT6DL2	BP022194.D	03 Oct 2024 13:48	Not Required	RC/JU	Not Ok
40	PB163716BS	SLCS716	BP022195.D	03 Oct 2024 14:28		RC/JU	Ok,M
41	P4016-11DL2	DCVW2DL2	BP022196.D	03 Oct 2024 15:09	Medium Level Extraction required, Run 10X dilution in Medium level	RC/JU	Dilution
42	SSTDCCC020	SSTD020785	BP022197.D	03 Oct 2024 16:30		RC/JU	Ok,M
43	PB163456BL	SBLK456	BP022198.D	03 Oct 2024 17:11		RC/JU	Ok,M
44	P4017-03DL	DDC42DL	BP022199.D	03 Oct 2024 17:52	Need Further 3X Dilution, and also Medium Level Extraction is Required	RC/JU	Dilution
45	P3927-08ME	DDC93ME	BP022200.D	03 Oct 2024 18:33	Need 10X Dilution	RC/JU	Dilution
46	P4018-03ME	DDCA1ME	BP022201.D	03 Oct 2024 19:13	Need 10X Dilution	RC/JU	Dilution
47	P4018-10ME	DDCB6ME	BP022202.D	03 Oct 2024 19:54	Need 10X Dilution	RC/JU	Dilution
48	P4018-15ME	DDCK3ME	BP022203.D	03 Oct 2024 20:35	Need 10X Dilution	RC/JU	Dilution
49	P4018-08ME	DDCB4ME	BP022204.D	03 Oct 2024 21:15	Need 10X Dilution	RC/JU	Dilution

Instrument ID: BNA_P

Daily Analysis Runlog For Sequence/QC Batch ID # BP100224

Review By	yogesh	Review On	10/4/2024 12:33:53 AM		
Supervise By	mohammad	Supervise On	10/5/2024 4:48:07 AM		
SubDirectory	BP100224	HP Acquire Method	BNA_P	HP Processing Method	BP092424
STD. NAME	STD REF.#				
Tune/Reschk	SP6633				
Initial Calibration Stds	SP6585,SP6580,SP6581,SP6582,SP6583,SP6584				
CCC	SP6581				
Internal Standard/PEM	S12317,10ul/1000ul sample				
ICV/I.BLK	SP6618				
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

50	P4018-09ME	DDCB5ME	BP022205.D	03 Oct 2024 21:56	Need 10X Dilution	RC/JU	Dilution
51	P4019-02ME	DDC38ME	BP022206.D	03 Oct 2024 22:37	Need 10X Dilution	RC/JU	Dilution
52	P4019-09ME	DDCE5ME	BP022207.D	03 Oct 2024 23:17	Need 10X Dilution	RC/JU	Dilution
53	P4019-22ME	DDCJ8ME	BP022208.D	03 Oct 2024 23:58	Need 20X Dilution	RC/JU	Dilution
54	P4017-06DL	DDC69DL	BP022209.D	04 Oct 2024 00:38		RC/JU	Ok
55	P4017-07DL	DDC70DL	BP022210.D	04 Oct 2024 01:19	Medium Level Extraction Required	RC/JU	Dilution
56	PB163860BS	SLCS860	BP022211.D	04 Oct 2024 02:00		RC/JU	Ok,M
57	SSTDCCC020	SSTD020786	BP022212.D	04 Oct 2024 03:21		RC/JU	Ok,M
58	PB163860BL	SBLK860	BP022213.D	04 Oct 2024 04:01		RC/JU	Ok,M
59	P4017-11DL	DDC86DL	BP022214.D	04 Oct 2024 04:42	Medium Level Extraction Required	RC/JU	Dilution
60	P4017-17DL	DDCD3DL	BP022215.D	04 Oct 2024 05:22	Need Further 3X Dilution	RC/JU	Dilution
61	P4017-17DL2	DDCD3DL2	BP022216.D	04 Oct 2024 06:02		RC/JU	Ok,M
62	P4017-14DL	DDC96DL	BP022217.D	04 Oct 2024 06:43	Medium Level Extraction Required	RC/JU	Dilution
63	P4017-20DL	DDCD6DL	BP022218.D	04 Oct 2024 07:24	Need Further 3X Dilution, and also Medium Level Extraction is Required	RC/JU	Dilution
64	P4017-02DL	DDC41DL	BP022219.D	04 Oct 2024 08:04	Medium Level Extraction Required	RC/JU	Dilution
65	P4017-19DL	DDCD5DL	BP022220.D	04 Oct 2024 08:45	Medium Level Extraction Required	RC/JU	Dilution
66	P4018-03MEDL	DDCA1MEDL	BP022221.D	04 Oct 2024 09:25	Need further 5X Dilution	RC/JU	Dilution

Instrument ID: BNA_P

Daily Analysis Runlog For Sequence/QC Batch ID # BP100224

Review By	yogesh	Review On	10/4/2024 12:33:53 AM		
Supervise By	mohammad	Supervise On	10/5/2024 4:48:07 AM		
SubDirectory	BP100224	HP Acquire Method	BNA_P	HP Processing Method	BP092424
STD. NAME		STD REF.#			
Tune/Reschk		SP6633			
Initial Calibration Stds		SP6585,SP6580,SP6581,SP6582,SP6583,SP6584			
CCC		SP6581			
Internal Standard/PEM		S12317,10ul/1000ul sample			
ICV/I.BLK		SP6618			
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

67	P4018-08MEDL	DDCB4MEDL	BP022222.D	04 Oct 2024 10:06		RC/JU	Ok,M
68	P4018-09MEDL	DDCB5MEDL	BP022223.D	04 Oct 2024 10:46		RC/JU	Ok,M
69	P4018-10MEDL	DDCB6MEDL	BP022224.D	04 Oct 2024 11:27		RC/JU	Ok,M
70	SSTDCCC020EC	SSTD020787	BP022225.D	04 Oct 2024 12:07		RC/JU	Ok,M

M : Manual Integration