

Report of Analysis

Client:	USEPA CLP Organics	Date Collected:	09/10/24
Project:	51486	Date Received:	09/14/24
Client Sample ID:	DDCA6	SDG No.:	P4022
Lab Sample ID:	P4022-08	Matrix:	Water
Analytical Method:	SFAM_SVOC	% Solid:	0
Sample Wt/Vol:	1000 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SPLP BNA Group1
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BP022117.D	1	09/26/24 08:20	09/30/24 18:03	PB163717

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
123-91-1	1,4-Dioxane	0.35	U	0.35	2.00	ug/L
123-91-1	1,4-Dioxane	0.35	U	0.35	2.00	ug/L
100-52-7	Benzaldehyde	1.10	U	1.10	10.0	ug/L
100-52-7	Benzaldehyde	1.10	U	1.10	10.0	ug/L
108-95-2	Phenol	1.60	U	1.60	10.0	ug/L
108-95-2	Phenol	1.60	U	1.60	10.0	ug/L
111-44-4	Bis(2-Chloroethyl)ether	1.40	U	1.40	10.0	ug/L
111-44-4	Bis(2-Chloroethyl)ether	1.40	U	1.40	10.0	ug/L
95-57-8	2-Chlorophenol	0.91	U	0.91	5.00	ug/L
95-57-8	2-Chlorophenol	0.91	U	0.91	5.00	ug/L
95-48-7	2-Methylphenol	1.50	U	1.50	10.0	ug/L
95-48-7	2-Methylphenol	1.50	U	1.50	10.0	ug/L
108-60-1	2,2-oxybis(1-Chloropropane)	1.90	U	1.90	10.0	ug/L
108-60-1	2,2-oxybis(1-Chloropropane)	1.90	U	1.90	10.0	ug/L
98-86-2	Acetophenone	1.50	U	1.50	10.0	ug/L
98-86-2	Acetophenone	1.50	U	1.50	10.0	ug/L
106-44-5	4-Methylphenol	1.70	U	1.70	10.0	ug/L
106-44-5	4-Methylphenol	1.70	U	1.70	10.0	ug/L
621-64-7	N-Nitroso-di-n-propylamine	1.90	U	1.90	5.00	ug/L
621-64-7	N-Nitroso-di-n-propylamine	1.90	U	1.90	5.00	ug/L
67-72-1	Hexachloroethane	1.10	U	1.10	5.00	ug/L
67-72-1	Hexachloroethane	1.10	U	1.10	5.00	ug/L
98-95-3	Nitrobenzene	1.40	U	1.40	5.00	ug/L
98-95-3	Nitrobenzene	1.40	U	1.40	5.00	ug/L
78-59-1	Isophorone	1.80	J	1.70	5.00	ug/L
78-59-1	Isophorone	1.80	J	1.70	5.00	ug/L
88-75-5	2-Nitrophenol	1.20	U	1.20	5.00	ug/L
88-75-5	2-Nitrophenol	1.20	U	1.20	5.00	ug/L
105-67-9	2,4-Dimethylphenol	0.69	U	0.69	5.00	ug/L

Report of Analysis

Client:	USEPA CLP Organics	Date Collected:	09/10/24
Project:	51486	Date Received:	09/14/24
Client Sample ID:	DDCA6	SDG No.:	P4022
Lab Sample ID:	P4022-08	Matrix:	Water
Analytical Method:	SFAM_SVOC	% Solid:	0
Sample Wt/Vol:	1000 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SPLP BNA Group1
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BP022117.D	1	09/26/24 08:20	09/30/24 18:03	PB163717

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
105-67-9	2,4-Dimethylphenol	0.69	U	0.69	5.00	ug/L
111-91-1	Bis(2-Chloroethoxy)methane	1.50	U	1.50	5.00	ug/L
111-91-1	Bis(2-Chloroethoxy)methane	1.50	U	1.50	5.00	ug/L
120-83-2	2,4-Dichlorophenol	1.10	U	1.10	5.00	ug/L
120-83-2	2,4-Dichlorophenol	1.10	U	1.10	5.00	ug/L
91-20-3	Naphthalene	9.70		0.94	5.00	ug/L
91-20-3	Naphthalene	9.70		0.94	5.00	ug/L
106-47-8	4-Chloroaniline	0.94	U	0.94	10.0	ug/L
106-47-8	4-Chloroaniline	0.94	U	0.94	10.0	ug/L
87-68-3	Hexachlorobutadiene	1.50	U	1.50	5.00	ug/L
87-68-3	Hexachlorobutadiene	1.50	U	1.50	5.00	ug/L
105-60-2	Caprolactam	2.60	U	2.60	10.0	ug/L
105-60-2	Caprolactam	2.60	U	2.60	10.0	ug/L
59-50-7	4-Chloro-3-methylphenol	1.60	U	1.60	5.00	ug/L
59-50-7	4-Chloro-3-methylphenol	1.60	U	1.60	5.00	ug/L
90-12-0	1-Methylnaphthalene	9.90		0.91	5.00	ug/L
90-12-0	1-Methylnaphthalene	9.90		0.91	5.00	ug/L
91-57-6	2-Methylnaphthalene	16.0		1.10	5.00	ug/L
91-57-6	2-Methylnaphthalene	16.0		1.10	5.00	ug/L
77-47-4	Hexachlorocyclopentadiene	3.10	U	3.10	10.0	ug/L
77-47-4	Hexachlorocyclopentadiene	3.10	U	3.10	10.0	ug/L
88-06-2	2,4,6-Trichlorophenol	2.00	U	2.00	5.00	ug/L
88-06-2	2,4,6-Trichlorophenol	2.00	U	2.00	5.00	ug/L
95-95-4	2,4,5-Trichlorophenol	1.60	U	1.60	5.00	ug/L
95-95-4	2,4,5-Trichlorophenol	1.60	U	1.60	5.00	ug/L
92-52-4	1,1-Biphenyl	1.10	J	1.10	5.00	ug/L
92-52-4	1,1-Biphenyl	1.10	J	1.10	5.00	ug/L
91-58-7	2-Chloronaphthalene	1.10	U	1.10	5.00	ug/L
91-58-7	2-Chloronaphthalene	1.10	U	1.10	5.00	ug/L
88-74-4	2-Nitroaniline	1.80	U	1.80	5.00	ug/L
88-74-4	2-Nitroaniline	1.80	U	1.80	5.00	ug/L

Report of Analysis

Client:	USEPA CLP Organics	Date Collected:	09/10/24
Project:	51486	Date Received:	09/14/24
Client Sample ID:	DDCA6	SDG No.:	P4022
Lab Sample ID:	P4022-08	Matrix:	Water
Analytical Method:	SFAM_SVOC	% Solid:	0
Sample Wt/Vol:	1000 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SPLP BNA Group1
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BP022117.D	1	09/26/24 08:20	09/30/24 18:03	PB163717

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
131-11-3	Dimethylphthalate	1.10	U	1.10	5.00	ug/L
131-11-3	Dimethylphthalate	1.10	U	1.10	5.00	ug/L
606-20-2	2,6-Dinitrotoluene	1.70	U	1.70	5.00	ug/L
606-20-2	2,6-Dinitrotoluene	1.70	U	1.70	5.00	ug/L
208-96-8	Acenaphthylene	1.20	U	1.20	5.00	ug/L
208-96-8	Acenaphthylene	1.20	U	1.20	5.00	ug/L
99-09-2	3-Nitroaniline	1.60	U	1.60	10.0	ug/L
99-09-2	3-Nitroaniline	1.60	U	1.60	10.0	ug/L
83-32-9	Acenaphthene	0.96	U	0.96	5.00	ug/L
83-32-9	Acenaphthene	0.96	U	0.96	5.00	ug/L
51-28-5	2,4-Dinitrophenol	1.80	U	1.80	10.0	ug/L
51-28-5	2,4-Dinitrophenol	1.80	U	1.80	10.0	ug/L
100-02-7	4-Nitrophenol	1.40	U	1.40	10.0	ug/L
100-02-7	4-Nitrophenol	1.40	U	1.40	10.0	ug/L
132-64-9	Dibenzofuran	1.10	U	1.10	5.00	ug/L
132-64-9	Dibenzofuran	1.10	U	1.10	5.00	ug/L
121-14-2	2,4-Dinitrotoluene	1.50	U	1.50	5.00	ug/L
121-14-2	2,4-Dinitrotoluene	1.50	U	1.50	5.00	ug/L
84-66-2	Diethylphthalate	1.30	U	1.30	5.00	ug/L
84-66-2	Diethylphthalate	1.30	U	1.30	5.00	ug/L
86-73-7	Fluorene	0.94	U	0.94	5.00	ug/L
86-73-7	Fluorene	0.94	U	0.94	5.00	ug/L
7005-72-3	4-Chlorophenyl-phenylether	1.10	U	1.10	5.00	ug/L
7005-72-3	4-Chlorophenyl-phenylether	1.10	U	1.10	5.00	ug/L
100-01-6	4-Nitroaniline	0.83	U	0.83	10.0	ug/L
100-01-6	4-Nitroaniline	0.83	U	0.83	10.0	ug/L
534-52-1	4,6-Dinitro-2-methylphenol	2.40	U	2.40	10.0	ug/L
534-52-1	4,6-Dinitro-2-methylphenol	2.40	U	2.40	10.0	ug/L
86-30-6	N-Nitrosodiphenylamine	1.30	U	1.30	5.00	ug/L
86-30-6	N-Nitrosodiphenylamine	1.30	U	1.30	5.00	ug/L
95-94-3	1,2,4,5-Tetrachlorobenzene	1.40	U	1.40	5.00	ug/L

Report of Analysis

Client:	USEPA CLP Organics	Date Collected:	09/10/24
Project:	51486	Date Received:	09/14/24
Client Sample ID:	DDCA6	SDG No.:	P4022
Lab Sample ID:	P4022-08	Matrix:	Water
Analytical Method:	SFAM_SVOC	% Solid:	0
Sample Wt/Vol:	1000 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SPLP BNA Group1
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BP022117.D	1	09/26/24 08:20	09/30/24 18:03	PB163717

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
95-94-3	1,2,4,5-Tetrachlorobenzene	1.40	U	1.40	5.00	ug/L
101-55-3	4-Bromophenyl-phenylether	0.95	U	0.95	5.00	ug/L
101-55-3	4-Bromophenyl-phenylether	0.95	U	0.95	5.00	ug/L
118-74-1	Hexachlorobenzene	1.20	U	1.20	5.00	ug/L
118-74-1	Hexachlorobenzene	1.20	U	1.20	5.00	ug/L
1912-24-9	Atrazine	1.80	U	1.80	10.0	ug/L
1912-24-9	Atrazine	1.80	U	1.80	10.0	ug/L
87-86-5	Pentachlorophenol	2200	E	1.90	10.0	ug/L
87-86-5	Pentachlorophenol	2200	E	1.90	10.0	ug/L
85-01-8	Phenanthrene	1.20	U	1.20	5.00	ug/L
85-01-8	Phenanthrene	1.20	U	1.20	5.00	ug/L
120-12-7	Anthracene	1.10	U	1.10	5.00	ug/L
120-12-7	Anthracene	1.10	U	1.10	5.00	ug/L
86-74-8	Carbazole	1.10	U	1.10	10.0	ug/L
86-74-8	Carbazole	1.10	U	1.10	10.0	ug/L
84-74-2	Di-n-butylphthalate	1.90	U	1.90	5.00	ug/L
84-74-2	Di-n-butylphthalate	1.90	U	1.90	5.00	ug/L
206-44-0	Fluoranthene	1.90	U	1.90	5.00	ug/L
206-44-0	Fluoranthene	1.90	U	1.90	5.00	ug/L
129-00-0	Pyrene	1.80	U	1.80	5.00	ug/L
129-00-0	Pyrene	1.80	U	1.80	5.00	ug/L
85-68-7	Butylbenzylphthalate	2.70	U	2.70	5.00	ug/L
85-68-7	Butylbenzylphthalate	2.70	U	2.70	5.00	ug/L
91-94-1	3,3-Dichlorobenzidine	1.40	U	1.40	10.0	ug/L
91-94-1	3,3-Dichlorobenzidine	1.40	U	1.40	10.0	ug/L
56-55-3	Benzo(a)anthracene	1.10	U	1.10	5.00	ug/L
56-55-3	Benzo(a)anthracene	1.10	U	1.10	5.00	ug/L
218-01-9	Chrysene	1.10	U	1.10	5.00	ug/L
218-01-9	Chrysene	1.10	U	1.10	5.00	ug/L
117-81-7	Bis(2-ethylhexyl)phthalate	3.00	U	3.00	5.00	ug/L
117-81-7	Bis(2-ethylhexyl)phthalate	3.00	U	3.00	5.00	ug/L

Report of Analysis

Client:	USEPA CLP Organics	Date Collected:	09/10/24
Project:	51486	Date Received:	09/14/24
Client Sample ID:	DDCA6	SDG No.:	P4022
Lab Sample ID:	P4022-08	Matrix:	Water
Analytical Method:	SFAM_SVOC	% Solid:	0
Sample Wt/Vol:	1000 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SPLP BNA Group1
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BP022117.D	1	09/26/24 08:20	09/30/24 18:03	PB163717

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
117-84-0	Di-n-octyl phthalate	2.40	U	2.40	10.0	ug/L
117-84-0	Di-n-octyl phthalate	2.40	U	2.40	10.0	ug/L
205-99-2	Benzo(b)fluoranthene	1.40	U	1.40	5.00	ug/L
205-99-2	Benzo(b)fluoranthene	1.40	U	1.40	5.00	ug/L
207-08-9	Benzo(k)fluoranthene	1.40	U	1.40	5.00	ug/L
207-08-9	Benzo(k)fluoranthene	1.40	U	1.40	5.00	ug/L
50-32-8	Benzo(a)pyrene	1.10	U	1.10	5.00	ug/L
50-32-8	Benzo(a)pyrene	1.10	U	1.10	5.00	ug/L
193-39-5	Indeno(1,2,3-cd)pyrene	1.40	U	1.40	5.00	ug/L
193-39-5	Indeno(1,2,3-cd)pyrene	1.40	U	1.40	5.00	ug/L
53-70-3	Dibenzo(a,h)anthracene	1.30	U	1.30	5.00	ug/L
53-70-3	Dibenzo(a,h)anthracene	1.30	U	1.30	5.00	ug/L
191-24-2	Benzo(g,h,i)perylene	1.20	U	1.20	5.00	ug/L
191-24-2	Benzo(g,h,i)perylene	1.20	U	1.20	5.00	ug/L
58-90-2	2,3,4,6-Tetrachlorophenol	370	E	1.50	5.00	ug/L
58-90-2	2,3,4,6-Tetrachlorophenol	370	E	1.50	5.00	ug/L

SURROGATES

17647-74-4	1,4-Dioxane-d8	3.70		15 - 120	46%	SPK: 8
17647-74-4	1,4-Dioxane-d8	3.70		15 - 120	46%	SPK: 8
7291-22-7	Pyridine-d5	4.80	*	20 - 120	12%	SPK: 40
7291-22-7	Pyridine-d5	4.80	*	20 - 120	12%	SPK: 40
4165-62-2	Phenol-d5	9.50		10 - 130	24%	SPK: 40
4165-62-2	Phenol-d5	9.50		10 - 130	24%	SPK: 40
93952-02-4	Bis-(2-Chloroethyl)ether-d8	21.4		25 - 120	54%	SPK: 40
93952-02-4	Bis-(2-Chloroethyl)ether-d8	21.4		25 - 120	54%	SPK: 40
93951-73-6	2-Chlorophenol-d4	23.1		20 - 130	58%	SPK: 40
93951-73-6	2-Chlorophenol-d4	23.1		20 - 130	58%	SPK: 40
190780-66-6	4-Methylphenol-d8	17.8		25 - 125	44%	SPK: 40
190780-66-6	4-Methylphenol-d8	17.8		25 - 125	44%	SPK: 40
4165-60-0	Nitrobenzene-d5	23.5		20 - 125	59%	SPK: 40

Report of Analysis

Client:	USEPA CLP Organics	Date Collected:	09/10/24
Project:	51486	Date Received:	09/14/24
Client Sample ID:	DDCA6	SDG No.:	P4022
Lab Sample ID:	P4022-08	Matrix:	Water
Analytical Method:	SFAM_SVOC	% Solid:	0
Sample Wt/Vol:	1000 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SPLP BNA Group1
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BP022117.D	1	09/26/24 08:20	09/30/24 18:03	PB163717

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
4165-60-0	Nitrobenzene-d5	23.5		20 - 125	59%	SPK: 40
93951-78-1	2-Nitrophenol-d4	24.5		20 - 130	61%	SPK: 40
93951-78-1	2-Nitrophenol-d4	24.5		20 - 130	61%	SPK: 40
93951-74-7	2,4-Dichlorophenol-d3	23.2		20 - 120	58%	SPK: 40
93951-74-7	2,4-Dichlorophenol-d3	23.2		20 - 120	58%	SPK: 40
191656-33-4	4-Chloroaniline-d4	3.10		1 - 146	8%	SPK: 40
191656-33-4	4-Chloroaniline-d4	3.10		1 - 146	8%	SPK: 40
85448-30-2	Dimethylphthalate-d6	27.4		25 - 130	69%	SPK: 40
85448-30-2	Dimethylphthalate-d6	27.4		25 - 130	69%	SPK: 40
93951-97-4	Acenaphthylene-d8	27.4		10 - 130	69%	SPK: 40
93951-97-4	Acenaphthylene-d8	27.4		10 - 130	69%	SPK: 40
93951-79-2	4-Nitrophenol-d4	9.50		10 - 150	24%	SPK: 40
93951-79-2	4-Nitrophenol-d4	9.50		10 - 150	24%	SPK: 40
81103-79-9	Fluorene-d10	26.8		25 - 125	67%	SPK: 40
81103-79-9	Fluorene-d10	26.8		25 - 125	67%	SPK: 40
93951-76-9	4,6-Dinitro-2-methylphenol-d2	28.6		10 - 130	71%	SPK: 40
93951-76-9	4,6-Dinitro-2-methylphenol-d2	28.6		10 - 130	71%	SPK: 40
1719-06-8	Anthracene-d10	28.7		25 - 130	72%	SPK: 40
1719-06-8	Anthracene-d10	28.7		25 - 130	72%	SPK: 40
1718-52-1	Pyrene-d10	28.6		15 - 130	72%	SPK: 40
1718-52-1	Pyrene-d10	28.6		15 - 130	72%	SPK: 40
63466-71-7	Benzo(a)pyrene-d12	28.6		20 - 130	71%	SPK: 40
63466-71-7	Benzo(a)pyrene-d12	28.6		20 - 130	71%	SPK: 40

INTERNAL STANDARDS

3855-82-1	1,4-Dichlorobenzene-d4	109000	7.699
3855-82-1	1,4-Dichlorobenzene-d4	109000	7.699
1146-65-2	Naphthalene-d8	404000	10.457
1146-65-2	Naphthalene-d8	404000	10.457
15067-26-2	Acenaphthene-d10	301000	14.304
15067-26-2	Acenaphthene-d10	301000	14.304

Report of Analysis

Client:	USEPA CLP Organics		Date Collected:	09/10/24
Project:	51486		Date Received:	09/14/24
Client Sample ID:	DDCA6		SDG No.:	P4022
Lab Sample ID:	P4022-08		Matrix:	Water
Analytical Method:	SFAM_SVOC		% Solid:	0
Sample Wt/Vol:	1000	Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:		uL	Test:	SPLP BNA Group1
Extraction Type :		Decanted : N	Level :	LOW
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BP022117.D	1	09/26/24 08:20	09/30/24 18:03	PB163717

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
1517-22-2	Phenanthrene-d10	706000	17.11			
1517-22-2	Phenanthrene-d10	706000	17.11			
1719-03-5	Chrysene-d12	802000	21.539			
1719-03-5	Chrysene-d12	802000	21.539			
1520-96-3	Perylene-d12	926000	24.821			
1520-96-3	Perylene-d12	926000	24.821			
TENTATIVE IDENTIFIED COMPOUNDS						
000123-42-2	2-Pentanone, 4-hydroxy-4-methyl-	7.90	A		4.86	ug/L
000123-42-2	2-Pentanone, 4-hydroxy-4-methyl-	7.90	A		4.86	ug/L
000108-38-3	Benzene, 1,3-dimethyl-	5.80	J		5.38	ug/L
000108-38-3	Benzene, 1,3-dimethyl-	5.80	J		5.38	ug/L
000123-05-7	Hexanal, 2-ethyl-	7.10	J		6.62	ug/L
000123-05-7	Hexanal, 2-ethyl-	7.10	J		6.62	ug/L
000611-14-3	Benzene, 1-ethyl-2-methyl-	23.0	J		6.80	ug/L
000611-14-3	Benzene, 1-ethyl-2-methyl-	23.0	J		6.80	ug/L
000526-73-8	Benzene, 1,2,3-trimethyl-	21.0	J		6.93	ug/L
000526-73-8	Benzene, 1,2,3-trimethyl-	21.0	J		6.93	ug/L
000622-96-8	Benzene, 1-ethyl-4-methyl-	22.0	J		7.10	ug/L
000622-96-8	Benzene, 1-ethyl-4-methyl-	22.0	J		7.10	ug/L
019549-80-5	2-Heptanone, 4,6-dimethyl-	27.0	J		7.14	ug/L
019549-80-5	2-Heptanone, 4,6-dimethyl-	27.0	J		7.14	ug/L
000108-67-8	Mesitylene	56.0	J		7.35	ug/L
000108-67-8	Mesitylene	56.0	J		7.35	ug/L
000104-76-7	1-Hexanol, 2-ethyl-	110	J		7.76	ug/L
000104-76-7	1-Hexanol, 2-ethyl-	110	J		7.76	ug/L
000095-63-6	Benzene, 1,2,4-trimethyl-	29.0	J		7.81	ug/L
000095-63-6	Benzene, 1,2,4-trimethyl-	29.0	J		7.81	ug/L
	unknown-01	10.0	J		7.93	ug/L
	unknown-01	10.0	J		7.93	ug/L
	unknown-02	12.0	J		8.17	ug/L

Report of Analysis

Client:	USEPA CLP Organics	Date Collected:	09/10/24
Project:	51486	Date Received:	09/14/24
Client Sample ID:	DDCA6	SDG No.:	P4022
Lab Sample ID:	P4022-08	Matrix:	Water
Analytical Method:	SFAM_SVOC	% Solid:	0
Sample Wt/Vol:	1000 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SPLP BNA Group1
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BP022117.D	1	09/26/24 08:20	09/30/24 18:03	PB163717

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
	unknown-02	12.0	J		8.17	ug/L
000934-74-7	Benzene, 1-ethyl-3,5-dimethyl-	20.0	J		8.33	ug/L
000934-74-7	Benzene, 1-ethyl-3,5-dimethyl-	20.0	J		8.33	ug/L
001074-55-1	Benzene, 1-methyl-4-propyl-	13.0	J		8.49	ug/L
001074-55-1	Benzene, 1-methyl-4-propyl-	13.0	J		8.49	ug/L
000933-98-2	Benzene, 1-ethyl-2,3-dimethyl-	13.0	J		8.63	ug/L
000933-98-2	Benzene, 1-ethyl-2,3-dimethyl-	13.0	J		8.63	ug/L
000527-84-4	o-Cymene	15.0	J		8.69	ug/L
000527-84-4	o-Cymene	15.0	J		8.69	ug/L
000149-57-5	Hexanoic acid, 2-ethyl-	11.0	J		9.00	ug/L
000149-57-5	Hexanoic acid, 2-ethyl-	11.0	J		9.00	ug/L
000504-20-1	Phorone	11.0	J		9.11	ug/L
000504-20-1	Phorone	11.0	J		9.11	ug/L
1000484-18-1	1,3-Hexanediol, 2-ethyl- (isomer 2	8.00	J		10.8	ug/L
1000484-18-1	1,3-Hexanediol, 2-ethyl- (isomer 2	8.00	J		10.8	ug/L
	unknown-03	12.0	J		10.8	ug/L
	unknown-03	12.0	J		10.8	ug/L
116436-59-0	3-Trifluoroacetoxy-6-ethyldecane	24.0	J		12.5	ug/L
116436-59-0	3-Trifluoroacetoxy-6-ethyldecane	24.0	J		12.5	ug/L
	unknown-04	9.70	J		13.4	ug/L
	unknown-04	9.70	J		13.4	ug/L
002349-07-7	Propanoic acid, 2-methyl-, hexyl e	9.70	J		13.6	ug/L
002349-07-7	Propanoic acid, 2-methyl-, hexyl e	9.70	J		13.6	ug/L
000529-16-8	Bicyclo[2.2.1]hept-2-ene, 2,3-dime	27.0	J		13.7	ug/L
000529-16-8	Bicyclo[2.2.1]hept-2-ene, 2,3-dime	27.0	J		13.7	ug/L
001502-22-3	2-(1-Cyclohexenyl)cyclohexanone	14.0	J		14.7	ug/L
001502-22-3	2-(1-Cyclohexenyl)cyclohexanone	14.0	J		14.7	ug/L
	unknown-05	6.90	J		14.8	ug/L
	unknown-05	6.90	J		14.8	ug/L
000935-95-5	Phenol, 2,3,5,6-tetrachloro-	7.30	J		14.9	ug/L
000935-95-5	Phenol, 2,3,5,6-tetrachloro-	7.30	J		14.9	ug/L

Report of Analysis

Client:	USEPA CLP Organics	Date Collected:	09/10/24
Project:	51486	Date Received:	09/14/24
Client Sample ID:	DDCA6	SDG No.:	P4022
Lab Sample ID:	P4022-08	Matrix:	Water
Analytical Method:	SFAM_SVOC	% Solid:	0
Sample Wt/Vol:	1000 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SPLP BNA Group1
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BP022117.D	1	09/26/24 08:20	09/30/24 18:03	PB163717

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
000829-26-5	Naphthalene, 2,3,6-trimethyl-	6.20	J		15.1	ug/L
000829-26-5	Naphthalene, 2,3,6-trimethyl-	6.20	J		15.1	ug/L
000119-61-9	Benzophenone	14.0	J		15.7	ug/L
000119-61-9	Benzophenone	14.0	J		15.7	ug/L
000057-10-3	n-Hexadecanoic acid	4.00	J		18.0	ug/L
000057-10-3	n-Hexadecanoic acid	4.00	J		18.0	ug/L
E966796	Total Alkanes	15.0			99.0	ug/L
E966796	Total Alkanes	15.0			99.0	ug/L

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products