

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF031025

Review By	anahy	Review On	3/11/2025 9:11:47 AM
Supervise By	Jagrut	Supervise On	3/11/2025 10:18:40 AM
SubDirectory	BF031025	HP Acquire Method	BNA_F
		HP Processing Method	bf031025

STD. NAME	STD REF.#
Tune/Reschk	SP6717
Initial Calibration Stds	SP6722,SP6723,SP6724,SP6725,SP6726,SP6727,SP6728,SP6729
CCC	SP6725
Internal Standard/PEM	S12653,10ul/1000ul sample
ICV/I.BLK	SP6686
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BF141896.D	10 Mar 2025 10:31		RC/JU	Ok
2	SSTDICC2.5	SSTDICC2.5	BF141897.D	10 Mar 2025 11:01		RC/JU	Ok
3	SSTDICC005	SSTDICC005	BF141898.D	10 Mar 2025 11:30	Compound#9,32,54,65 removed from 5 ppm	RC/JU	Ok
4	SSTDICC010	SSTDICC010	BF141899.D	10 Mar 2025 12:00		RC/JU	Ok
5	SSTDICC020	SSTDICC020	BF141900.D	10 Mar 2025 12:29	Compound #54 Kept on LR, Method is good for DOD . Method failed for compound #77.	RC/JU	Ok
6	SSTDICCC040	SSTDICCC040	BF141901.D	10 Mar 2025 12:58		RC/JU	Ok
7	SSTDICC050	SSTDICC050	BF141902.D	10 Mar 2025 13:28	Not used	RC/JU	Not Ok
8	SSTDICC060	SSTDICC060	BF141903.D	10 Mar 2025 13:57	Compound#69 Removed from 60 ppm	RC/JU	Ok
9	SSTDICC080	SSTDICC080	BF141904.D	10 Mar 2025 14:27	Compound#9,69 Removed from 80 ppm	RC/JU	Ok,M
10	SSTDICC050	SSTDICC050	BF141905.D	10 Mar 2025 15:20		RC/JU	Ok
11	SSTDICV040	ICVBF031025	BF141906.D	10 Mar 2025 15:53		RC/JU	Ok
12	PB167012BL	PB167012BL	BF141907.D	10 Mar 2025 17:09		RC/JU	Ok
13	SP6752	SP6752	BF141908.D	10 Mar 2025 17:39	8270 SPIKE-SP6752	RC/JU	Ok,M

M : Manual Integration

Instrument ID: BNA_P

Daily Analysis Runlog For Sequence/QC Batch ID # BP093024

Review By	Jagrut	Review On	10/1/2024 11:04:18 AM
Supervise By	mohammad	Supervise On	10/2/2024 9:04:47 AM
SubDirectory	BP093024	HP Acquire Method	BNA_P
		HP Processing Method	BP092424

STD. NAME	STD REF.#
Tune/Reschk	SP6633
Initial Calibration Stds	SP6585,SP6580,SP6581,SP6582,SP6583,SP6584
CCC	SP6581
Internal Standard/PEM	S12317,10ul/1000ul sample
ICV/I.BLK	SP6618
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP762	DFTPP762	BP022111.D	30 Sep 2024 13:57		RC/JU	Ok
2	SSTDCCC020	SSTD020777	BP022112.D	30 Sep 2024 14:38		RC/JU	Ok,M
3	PB163717BL	SBLK717	BP022113.D	30 Sep 2024 15:19		RC/JU	Ok,M
4	PB163716BL	SBLK716	BP022114.D	30 Sep 2024 16:00		RC/JU	Ok
5	P4022-06	DDCD6	BP022115.D	30 Sep 2024 16:41	Need 5X & 50X Dilution	RC/JU	Dilution
6	P4022-07	DDCA1	BP022116.D	30 Sep 2024 17:22	Need 5X & 50X Dilution	RC/JU	Dilution
7	P4022-08	DDCA6	BP022117.D	30 Sep 2024 18:03	Need 50X Dilution	RC/JU	Dilution
8	P4022-01	DDC37	BP022118.D	30 Sep 2024 18:44	Need 50X Dilution	RC/JU	Dilution
9	P4022-02	DDC70	BP022119.D	30 Sep 2024 19:24	Need 20X Dilution	RC/JU	Dilution
10	P4022-03MS	DDC70MS	BP022120.D	30 Sep 2024 20:05	Internal Standard Fail	RC/JU	Not Ok
11	P4022-04MSD	DDC70MSD	BP022121.D	30 Sep 2024 20:46		RC/JU	Ok,M
12	PB163716TB	SLEB650	BP022122.D	30 Sep 2024 21:26		RC/JU	Ok
13	P4022-05	DDC95	BP022123.D	30 Sep 2024 22:07	Need 30X Dilution	RC/JU	Dilution
14	PB163716BS	SLCS716	BP022124.D	30 Sep 2024 22:48	Internal Standard Fail	RC/JU	Not Ok
15	PB163717BS	SLCS717	BP022125.D	30 Sep 2024 23:29		RC/JU	Ok,M
16	SSTDCCC020EC	SSTD020778	BP022126.D	01 Oct 2024 00:50		RC/JU	Ok,M
17	PB163565BL	SBLK565	BP022127.D	01 Oct 2024 01:31	END CCC Missing	RC/JU	Not Ok

M : Manual Integration

Instrument ID: BNA_P

Daily Analysis Runlog For Sequence/QC Batch ID # BP100124

Review By	yogesh	Review On	10/2/2024 8:42:48 AM
Supervise By	mohammad	Supervise On	10/4/2024 1:06:48 AM
SubDirectory	BP100124	HP Acquire Method	BNA_P
		HP Processing Method	BP092424

STD. NAME	STD REF.#
Tune/Reschk	SP6633
Initial Calibration Stds	SP6585,SP6580,SP6581,SP6582,SP6583,SP6584
CCC	SP6581
Internal Standard/PEM	S12317,10ul/1000ul sample
ICV/I.BLK	SP6618
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP763	DFTPP763	BP022128.D	01 Oct 2024 09:46		RC/JU	Ok
2	SSTDCCC020	SSTD020779	BP022129.D	01 Oct 2024 10:26		RC/JU	Ok,M
3	PB163565BL	SBLK565	BP022130.D	01 Oct 2024 11:06		RC/JU	Ok
4	P4010-01	C0AT8	BP022131.D	01 Oct 2024 11:47		RC/JU	Ok
5	P4010-02	C0AW6	BP022132.D	01 Oct 2024 12:27		RC/JU	Ok,M
6	P4010-05	C0AT9	BP022133.D	01 Oct 2024 13:08		RC/JU	Ok
7	P4010-06	C0AW0	BP022134.D	01 Oct 2024 13:48	Surrogate Fail	RC/JU	ReRun
8	P4010-07	C0AW1	BP022135.D	01 Oct 2024 14:29	DUP	RC/JU	Ok
9	P4010-08	C0AW9	BP022136.D	01 Oct 2024 15:10	Need 5X Dilution	RC/JU	Dilution
10	P4010-09	C0AX0	BP022137.D	01 Oct 2024 15:50		RC/JU	Ok,M
11	P4021-01	DDCG1	BP022138.D	01 Oct 2024 16:31		RC/JU	Ok
12	P4021-02MS	DDCG1MS	BP022139.D	01 Oct 2024 17:12		RC/JU	Ok,M
13	P4021-03MSD	DDCG1MSD	BP022140.D	01 Oct 2024 17:53		RC/JU	Ok,M
14	PB163565BS	SLCS565	BP022141.D	01 Oct 2024 18:34		RC/JU	Ok,M
15	SSTDCCC020	SSTD020780	BP022142.D	01 Oct 2024 19:55		RC/JU	Ok
16	PB163717BL	SBLK717	BP022143.D	01 Oct 2024 20:36	CCC Fail	RC/JU	Not Ok
17	P4010-06	C0AW0	BP022144.D	01 Oct 2024 21:17	CCC Fail	RC/JU	ReRun
18	P4022-06DL	DDCD6DL	BP022145.D	01 Oct 2024 21:58	Need Further 10X Dilution.	RC/JU	Dilution

Instrument ID: BNA_P

Daily Analysis Runlog For Sequence/QC Batch ID # BP100124

Review By	yogesh	Review On	10/2/2024 8:42:48 AM		
Supervise By	mohammad	Supervise On	10/4/2024 1:06:48 AM		
SubDirectory	BP100124	HP Acquire Method	BNA_P	HP Processing Method	BP092424
STD. NAME		STD REF.#			
Tune/Reschk		SP6633			
Initial Calibration Stds		SP6585,SP6580,SP6581,SP6582,SP6583,SP6584			
CCC		SP6581			
Internal Standard/PEM		S12317,10ul/1000ul sample			
ICV/I.BLK		SP6618			
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

19	P4022-06DL2	DDCD6DL2	BP022146.D	01 Oct 2024 22:38		RC/JU	Ok,M
20	P4022-07DL	DDCA1DL	BP022147.D	01 Oct 2024 23:19	Need Further 10X Dilution.	RC/JU	Dilution
21	P4022-07DL2	DDCA1DL2	BP022148.D	01 Oct 2024 23:59		RC/JU	Ok,M
22	P4022-08DL	DDCA6DL	BP022149.D	02 Oct 2024 00:40		RC/JU	Ok,M
23	P4022-01DL	DDC37DL	BP022150.D	02 Oct 2024 01:21	Not Match Dilution	RC/JU	Not Ok
24	P4022-02DL	DDC70DL	BP022151.D	02 Oct 2024 02:01		RC/JU	Ok,M
25	P4022-03MS	DDC70MS	BP022152.D	02 Oct 2024 02:42	CCC Fail	RC/JU	Not Ok
26	P4022-05DL	DDC95DL	BP022153.D	02 Oct 2024 03:22		RC/JU	Ok,M
27	PB163716BS	SLCS716	BP022154.D	02 Oct 2024 04:03	CCC Fail	RC/JU	Not Ok
28	SSTDCCC020EC	SSTD020781	BP022155.D	02 Oct 2024 05:24		RC/JU	Ok,M

M : Manual Integration

Instrument ID: BNA_P

Daily Analysis Runlog For Sequence/QC Batch ID # BP100224

Review By	yogesh	Review On	10/4/2024 12:33:53 AM
Supervise By	mohammad	Supervise On	10/5/2024 4:48:07 AM
SubDirectory	BP100224	HP Acquire Method	BNA_P
		HP Processing Method	BP092424

STD. NAME	STD REF.#
Tune/Reschk	SP6633
Initial Calibration Stds	SP6585,SP6580,SP6581,SP6582,SP6583,SP6584
CCC	SP6581
Internal Standard/PEM	S12317,10ul/1000ul sample
ICV/I.BLK	SP6618
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP764	DFTPP764	BP022156.D	02 Oct 2024 10:41		RC/JU	Ok
2	SSTDCCC020	SSTD020782	BP022157.D	02 Oct 2024 11:23		RC/JU	Ok,M
3	PB163456BL	SBLK456	BP022158.D	02 Oct 2024 12:04	Analyzed for contamination check	RC/JU	Not Ok
4	P4016-11	DCVW2	BP022159.D	02 Oct 2024 12:44	Need 10X & 30X dilution & Medium level extraction required	RC/JU	Dilution
5	P4016-12	DCVW3	BP022160.D	02 Oct 2024 13:25	Need 30X dilution	RC/JU	Dilution
6	P4016-01	DCVT2	BP022161.D	02 Oct 2024 14:06		RC/JU	Ok
7	P4016-02	DCVT3	BP022162.D	02 Oct 2024 14:47		RC/JU	Ok
8	P4016-03	DCVT4	BP022163.D	02 Oct 2024 15:27		RC/JU	Ok
9	P4016-04	DCVT5	BP022164.D	02 Oct 2024 16:08		RC/JU	Ok
10	P4010-06	C0AW0	BP022165.D	02 Oct 2024 16:49		RC/JU	Ok,M
11	P4016-05	DCVT6	BP022166.D	02 Oct 2024 17:29	Need 10X	RC/JU	Dilution
12	P4016-06	DCVT7	BP022167.D	02 Oct 2024 18:10		RC/JU	Ok
13	P4016-07	DCVT8	BP022168.D	02 Oct 2024 18:51		RC/JU	Ok
14	P4016-08	DCVT9	BP022169.D	02 Oct 2024 19:32		RC/JU	Ok
15	SSTDCCC020	SSTD020783	BP022170.D	02 Oct 2024 20:53		RC/JU	Ok,M
16	PB163489BL	SBLK489	BP022171.D	02 Oct 2024 21:34		RC/JU	Ok
17	P4016-09	DCVW0	BP022172.D	02 Oct 2024 22:14		RC/JU	Ok

Instrument ID: BNA_P

Daily Analysis Runlog For Sequence/QC Batch ID # BP100224

Review By	yogesh	Review On	10/4/2024 12:33:53 AM		
Supervise By	mohammad	Supervise On	10/5/2024 4:48:07 AM		
SubDirectory	BP100224	HP Acquire Method	BNA_P	HP Processing Method	BP092424
STD. NAME		STD REF.#			
Tune/Reschk		SP6633			
Initial Calibration Stds		SP6585,SP6580,SP6581,SP6582,SP6583,SP6584			
CCC		SP6581			
Internal Standard/PEM		S12317,10ul/1000ul sample			
ICV/I.BLK		SP6618			
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

18	P4016-10	DCVW1	BP022173.D	02 Oct 2024 22:55		RC/JU	Ok
19	P4010-08DL	C0AW9DL	BP022174.D	02 Oct 2024 23:36		RC/JU	Ok,M
20	P4022-01DL	DDC37DL	BP022175.D	03 Oct 2024 00:16		RC/JU	Ok,M
21	P4017-03	DDC42	BP022176.D	03 Oct 2024 00:57	Need 10X Dilution and Medium level Extraction Required	RC/JU	Dilution
22	P4017-06	DDC69	BP022177.D	03 Oct 2024 01:37	Need 30X Dilution	RC/JU	Dilution
23	P4017-07	DDC70	BP022178.D	03 Oct 2024 02:18	Need 30X dilution and also Medium level extraction Required	RC/JU	Dilution
24	P4017-08MS	DDC70MS	BP022179.D	03 Oct 2024 02:58		RC/JU	Ok,M
25	P4017-09MSD	DDC70MSD	BP022180.D	03 Oct 2024 03:39		RC/JU	Ok,M
26	P4017-11	DDC86	BP022181.D	03 Oct 2024 04:20	Need 30X dilution and also Medium level extraction Required	RC/JU	Dilution
27	P4017-17	DDCD3	BP022182.D	03 Oct 2024 05:00	EPA COMMUNICATIOIN FOR Surrogate Fail, Need 10X & 30X Dilution	RC/JU	Dilution
28	PB163489BS	SLCS489	BP022183.D	03 Oct 2024 05:41		RC/JU	Ok,M
29	SSTDCCC020	SSTD020784	BP022184.D	03 Oct 2024 07:01		RC/JU	Ok
30	PB163456BL	SBLK456	BP022185.D	03 Oct 2024 07:42	Analyzed for contamination check	RC/JU	Ok
31	P4017-14	DDC96	BP022186.D	03 Oct 2024 08:23	Need 20X Dilution and Medium level Extraction Required.	RC/JU	Dilution
32	P4017-20	DDCD6	BP022187.D	03 Oct 2024 09:04	Need 10X Dilution and Medium level Extraction Required	RC/JU	Dilution
33	P4017-02	DDC41	BP022188.D	03 Oct 2024 09:44	Need 20X Dilution. Medium level Extraction Required	RC/JU	Dilution

Instrument ID: BNA_P

Daily Analysis Runlog For Sequence/QC Batch ID # BP100224

Review By	yogesh	Review On	10/4/2024 12:33:53 AM		
Supervise By	mohammad	Supervise On	10/5/2024 4:48:07 AM		
SubDirectory	BP100224	HP Acquire Method	BNA_P	HP Processing Method	BP092424
STD. NAME	STD REF.#				
Tune/Reschk	SP6633				
Initial Calibration Stds	SP6585,SP6580,SP6581,SP6582,SP6583,SP6584				
CCC	SP6581				
Internal Standard/PEM	S12317,10ul/1000ul sample				
ICV/I.BLK	SP6618				
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

34	P4017-19	DDCD5	BP022189.D	03 Oct 2024 10:25	Need 10X Dilution and Medium level Extraction Required	RC/JU	Dilution
35	P4016-11DL	DCVW2DL	BP022190.D	03 Oct 2024 11:05	Medium Level Extraction required, Run 10X dilution in Medium level	RC/JU	Dilution
36	P4016-11DL2	DCVW2DL2	BP022191.D	03 Oct 2024 11:46	Internal Standard Fail	RC/JU	Not Ok
37	P4016-12DL	DCVW3DL	BP022192.D	03 Oct 2024 12:27	Medium Level Extraction required, Run 10X dilution in Medium level	RC/JU	Dilution
38	P4016-05DL	DCVT6DL	BP022193.D	03 Oct 2024 13:07		RC/JU	Ok
39	P4016-05DL2	DCVT6DL2	BP022194.D	03 Oct 2024 13:48	Not Required	RC/JU	Not Ok
40	PB163716BS	SLCS716	BP022195.D	03 Oct 2024 14:28		RC/JU	Ok,M
41	P4016-11DL2	DCVW2DL2	BP022196.D	03 Oct 2024 15:09	Medium Level Extraction required, Run 10X dilution in Medium level	RC/JU	Dilution
42	SSTDCCC020	SSTD020785	BP022197.D	03 Oct 2024 16:30		RC/JU	Ok,M
43	PB163456BL	SBLK456	BP022198.D	03 Oct 2024 17:11		RC/JU	Ok,M
44	P4017-03DL	DDC42DL	BP022199.D	03 Oct 2024 17:52	Need Further 3X Dilution, and also Medium Level Extraction is Required	RC/JU	Dilution
45	P3927-08ME	DDC93ME	BP022200.D	03 Oct 2024 18:33	Need 10X Dilution	RC/JU	Dilution
46	P4018-03ME	DDCA1ME	BP022201.D	03 Oct 2024 19:13	Need 10X Dilution	RC/JU	Dilution
47	P4018-10ME	DDCB6ME	BP022202.D	03 Oct 2024 19:54	Need 10X Dilution	RC/JU	Dilution
48	P4018-15ME	DDCK3ME	BP022203.D	03 Oct 2024 20:35	Need 10X Dilution	RC/JU	Dilution
49	P4018-08ME	DDCB4ME	BP022204.D	03 Oct 2024 21:15	Need 10X Dilution	RC/JU	Dilution

Instrument ID: BNA_P

Daily Analysis Runlog For Sequence/QC Batch ID # BP100224

Review By	yogesh	Review On	10/4/2024 12:33:53 AM		
Supervise By	mohammad	Supervise On	10/5/2024 4:48:07 AM		
SubDirectory	BP100224	HP Acquire Method	BNA_P	HP Processing Method	BP092424
STD. NAME	STD REF.#				
Tune/Reschk	SP6633				
Initial Calibration Stds	SP6585,SP6580,SP6581,SP6582,SP6583,SP6584				
CCC	SP6581				
Internal Standard/PEM	S12317,10ul/1000ul sample				
ICV/I.BLK	SP6618				
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

50	P4018-09ME	DDCB5ME	BP022205.D	03 Oct 2024 21:56	Need 10X Dilution	RC/JU	Dilution
51	P4019-02ME	DDC38ME	BP022206.D	03 Oct 2024 22:37	Need 10X Dilution	RC/JU	Dilution
52	P4019-09ME	DDCE5ME	BP022207.D	03 Oct 2024 23:17	Need 10X Dilution	RC/JU	Dilution
53	P4019-22ME	DDCJ8ME	BP022208.D	03 Oct 2024 23:58	Need 20X Dilution	RC/JU	Dilution
54	P4017-06DL	DDC69DL	BP022209.D	04 Oct 2024 00:38		RC/JU	Ok
55	P4017-07DL	DDC70DL	BP022210.D	04 Oct 2024 01:19	Medium Level Extraction Required	RC/JU	Dilution
56	PB163860BS	SLCS860	BP022211.D	04 Oct 2024 02:00		RC/JU	Ok,M
57	SSTDCCC020	SSTD020786	BP022212.D	04 Oct 2024 03:21		RC/JU	Ok,M
58	PB163860BL	SBLK860	BP022213.D	04 Oct 2024 04:01		RC/JU	Ok,M
59	P4017-11DL	DDC86DL	BP022214.D	04 Oct 2024 04:42	Medium Level Extraction Required	RC/JU	Dilution
60	P4017-17DL	DDCD3DL	BP022215.D	04 Oct 2024 05:22	Need Further 3X Dilution	RC/JU	Dilution
61	P4017-17DL2	DDCD3DL2	BP022216.D	04 Oct 2024 06:02		RC/JU	Ok,M
62	P4017-14DL	DDC96DL	BP022217.D	04 Oct 2024 06:43	Medium Level Extraction Required	RC/JU	Dilution
63	P4017-20DL	DDCD6DL	BP022218.D	04 Oct 2024 07:24	Need Further 3X Dilution, and also Medium Level Extraction is Required	RC/JU	Dilution
64	P4017-02DL	DDC41DL	BP022219.D	04 Oct 2024 08:04	Medium Level Extraction Required	RC/JU	Dilution
65	P4017-19DL	DDCD5DL	BP022220.D	04 Oct 2024 08:45	Medium Level Extraction Required	RC/JU	Dilution
66	P4018-03MEDL	DDCA1MEDL	BP022221.D	04 Oct 2024 09:25	Need further 5X Dilution	RC/JU	Dilution

Instrument ID: BNA_P

Daily Analysis Runlog For Sequence/QC Batch ID # BP100224

Review By	yogesh	Review On	10/4/2024 12:33:53 AM		
Supervise By	mohammad	Supervise On	10/5/2024 4:48:07 AM		
SubDirectory	BP100224	HP Acquire Method	BNA_P	HP Processing Method	BP092424
STD. NAME		STD REF.#			
Tune/Reschk	SP6633				
Initial Calibration Stds	SP6585,SP6580,SP6581,SP6582,SP6583,SP6584				
CCC	SP6581				
Internal Standard/PEM	S12317,10ul/1000ul sample				
ICV/I.BLK	SP6618				
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

67	P4018-08MEDL	DDCB4MEDL	BP022222.D	04 Oct 2024 10:06		RC/JU	Ok,M
68	P4018-09MEDL	DDCB5MEDL	BP022223.D	04 Oct 2024 10:46		RC/JU	Ok,M
69	P4018-10MEDL	DDCB6MEDL	BP022224.D	04 Oct 2024 11:27		RC/JU	Ok,M
70	SSTDCCC020EC	SSTD020787	BP022225.D	04 Oct 2024 12:07		RC/JU	Ok,M

M : Manual Integration

Instrument ID: BNA_P

Daily Analysis Runlog For Sequence/QC Batch ID # BP100424

Review By	yogesh	Review On	10/5/2024 2:59:11 AM
Supervise By	mohammad	Supervise On	10/6/2024 1:18:02 AM
SubDirectory	BP100424	HP Acquire Method	BNA_P
		HP Processing Method	BP092424

STD. NAME	STD REF.#
Tune/Reschk	SP6633
Initial Calibration Stds	SP6585,SP6580,SP6581,SP6582,SP6583,SP6584
CCC	SP6581
Internal Standard/PEM	S12319,10ul/1000ul sample
ICV/I.BLK	SP6618
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP765	DFTPP765	BP022226.D	04 Oct 2024 13:00		RC/JU	Ok
2	SSTDCCC020	SSTD020788	BP022227.D	04 Oct 2024 13:41		RC/JU	Ok,M
3	PB163610BL	SBLK610	BP022228.D	04 Oct 2024 14:21		RC/JU	Ok
4	P4018-15MEDL	DDCK3MEDL	BP022229.D	04 Oct 2024 15:02		RC/JU	Ok
5	P3927-08MEDL	DDC93MEDL	BP022230.D	04 Oct 2024 15:43		RC/JU	Ok,M
6	P4019-02MEDL	DDC38MEDL	BP022231.D	04 Oct 2024 16:24		RC/JU	Ok
7	P4019-09MEDL	DDCE5MEDL	BP022232.D	04 Oct 2024 17:05		RC/JU	Ok
8	P4019-22MEDL	DDCJ8MEDL	BP022233.D	04 Oct 2024 17:46		RC/JU	Ok,M
9	P4017-03DL2	DDC42DL2	BP022234.D	04 Oct 2024 18:27	Medium Level Extraction Required	RC/JU	Dilution
10	P4017-20DL2	DDCD6DL2	BP022235.D	04 Oct 2024 19:08	Medium Level Extraction Required	RC/JU	Dilution
11	P4022-03MS	DDC70MS	BP022236.D	04 Oct 2024 19:48		RC/JU	Ok,M
12	P4083-03	A4ES3	BP022237.D	04 Oct 2024 20:29		RC/JU	Ok
13	P4083-08	A4EK3	BP022238.D	04 Oct 2024 21:10	Surrogate Fail	RC/JU	ReRun
14	PB163610BS	SLCS610	BP022239.D	04 Oct 2024 21:50		RC/JU	Ok,M
15	SSTDCCC020	SSTD020789	BP022240.D	04 Oct 2024 23:12		RC/JU	Ok,M
16	PB163635BL	SBLK635	BP022241.D	04 Oct 2024 23:54		RC/JU	Ok
17	PB163912BL	SBLK912	BP022242.D	05 Oct 2024 00:35		RC/JU	Ok

Instrument ID: BNA_P

Daily Analysis Runlog For Sequence/QC Batch ID # BP100424

Review By	yogesh	Review On	10/5/2024 2:59:11 AM		
Supervise By	mohammad	Supervise On	10/6/2024 1:18:02 AM		
SubDirectory	BP100424	HP Acquire Method	BNA_P	HP Processing Method	BP092424
STD. NAME		STD REF.#			
Tune/Reschk		SP6633			
Initial Calibration Stds		SP6585,SP6580,SP6581,SP6582,SP6583,SP6584			
CCC		SP6581			
Internal Standard/PEM		S12319,10ul/1000ul sample			
ICV/I.BLK		SP6618			
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

18	P4018-03MEDL2	DDCA1MEDL2	BP022243.D	05 Oct 2024 01:16		RC/JU	Ok,M
19	P4264-12	DCYO1	BP022244.D	05 Oct 2024 01:58	EB	RC/JU	Ok
20	P4237-03	A4EB9	BP022245.D	05 Oct 2024 02:39	EB	RC/JU	Ok
21	P4264-13	DCYO2	BP022246.D	05 Oct 2024 03:20	FB	RC/JU	Ok,M
22	P4083-02	A4ES2	BP022247.D	05 Oct 2024 04:01		RC/JU	Ok,M
23	P4083-07	A4EK2	BP022248.D	05 Oct 2024 04:43	Surrogate Fail	RC/JU	ReRun
24	P4083-09	A4EK4	BP022249.D	05 Oct 2024 05:24		RC/JU	Ok,M
25	P4083-10	A4EK5	BP022250.D	05 Oct 2024 06:05		RC/JU	Ok,M
26	P4083-11	A4EK9	BP022251.D	05 Oct 2024 06:45		RC/JU	Ok,M
27	P4083-13	A4EL3	BP022252.D	05 Oct 2024 07:26		RC/JU	Ok,M
28	P4284-10	A4EB3	BP022253.D	05 Oct 2024 08:07	EB	RC/JU	Ok
29	PB163912BS	SLCS912	BP022254.D	05 Oct 2024 08:48		RC/JU	Ok,M
30	SSTDCCC020EC	SSTD020790	BP022255.D	05 Oct 2024 10:09		RC/JU	Ok,M

M : Manual Integration