



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
 Fax : 908 789 8922

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH
 Lab Code: CHEM Case No.: P4022 SAS No.: P4022 SDG NO.: P4022
 EPA Sample No.: SSTD020783 Date Analyzed: 10/02/2024
 Lab File ID: BP022170.D Time Analyzed: 20:53
 Instrument ID: BNA_P GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	120788	7.699	470061	10.46	367777	14.31
UPPER LIMIT	241576	8.199	940122	10.958	735554	14.81
LOWER LIMIT	60394	7.199	235031	9.958	183889	13.81
EPA SAMPLE NO.						
01 DDC37DL	98858	7.70	393100	10.46	324409	14.30

IS1 (DCB) = 1,4-Dichlorobenzene-d4
 IS2 (NPT) = Naphthalene-d8
 IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.