

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: P4022 SAS No.: P4022 SDG NO.: P4022

EPA Sample No.: SSTD020783 Date Analyzed: 10/02/2024

Lab File ID: BP022170.D Time Analyzed: 20:53

Instrument ID: BNA_P GC Column: ZB-GR ID: 0.25 (mm)

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	860058	17.116	948490	21.551	1113280	24.839
UPPER LIMIT	1720120	17.616	1896980	22.051	2226560	25.339
LOWER LIMIT	430029	16.616	474245	21.051	556640	24.339
EPA SAMPLE NO.						
01 DDC37DL	807941	17.09	890992	21.53	1018190	24.82

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.