

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: P4022 SAS No.: P4022 SDG NO.: P4022

EPA Sample No.: SSTD020788 Date Analyzed: 10/04/2024

Lab File ID: BP022227.D Time Analyzed: 13:41

Instrument ID: BNA_P GC Column: ZB-GR ID: 0.25 (mm)

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	963201	17.092	1056950	21.539	1228930	24.833
UPPER LIMIT	1926400	17.592	2113900	22.039	2457860	25.333
LOWER LIMIT	481601	16.592	528475	21.039	614465	24.333
EPA SAMPLE NO.						
01 DDC70MS	888519	17.12	972771	21.53	1153720	24.82

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.