



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CASE NARRATIVE

ENTACT

Project Name: North Point

Project # N/A

Chemtech Project # P4103

Test Name: Metals ICP-Group1,Mercury

A. Number of Samples and Date of Receipt:

18 Solid samples were received on 09/18/2024.

B. Parameters:

According to the Chain of Custody document, the following analyses were requested: ASTM Ammonia, ASTM COD, ASTM Leach Extraction, ASTM Oil and Grease, ASTM TS, Corrosivity, Cyanide, Diesel Range Organics, EPH, Gasoline Range Organics, Hexavalent Chromium, Ignitability, Mercury, Metals Group1, Metals ICP-Group1, Paint Filter, PCB, RCRA CHARACTERISTICS, Reactive Cyanide, Reactive Sulfide, SVOC-TCL BNA -20, TCLP BNA, TCLP Extraction, TCLP Herbicide, TCLP ICP Metals, TCLP Mercury, TCLP Pesticide, TCLP VOA, TCLP ZHE Extraction, TCLP-FULL, TCLPMetals Group2, TS, TVS and VOC-TCLVOA-10. This data package contains results for Metals ICP-Group1,Mercury.

C. Analytical Techniques:

The analysis of Metals ICP-Group1 was based on method 6010D, digestion based on method 3050 (soils). The analysis and digestion of Mercury was based on method 7471B.

D. QA/ QC Samples:

The Holding Times were met for all analysis.

The Blank Spike met requirements for all samples.

The Duplicate analysis met criteria for all samples.

The Matrix Spike (NB-08-09192024MS) analysis met criteria for all samples except for Boron due to Chemical interference during Digestion Process.

Matrix Spike Duplicate (NB-08-09192024MSD) analysis met criteria for all samples except for Boron and Manganese due to Chemical interference during Digestion Process.

The Blank analysis did not indicate the presence of lab contamination.

The Calibration met the requirements.

The Serial Dilution met the acceptable requirements.

E. Additional Comments:



I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature_____