

Initial and Continuing Calibration Verification

Client: ENTACT

SDG No.: P4103

Project: North Point

RunNo.: LB132516

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1 Reactive Cyanide	mg/L	0.095	0.099	96	85-115	09/19/2024
Sample ID: CCV1 Reactive Cyanide	mg/L	0.24	0.25	96	90-110	09/19/2024
Sample ID: CCV2 Reactive Cyanide	mg/L	0.24	0.25	96	90-110	09/19/2024
Sample ID: CCV3 Reactive Cyanide	mg/L	0.24	0.25	96	90-110	09/19/2024
Sample ID: CCV4 Reactive Cyanide	mg/L	0.25	0.25	100	90-110	09/19/2024

Initial and Continuing Calibration Verification

Client: ENTACT

SDG No.: P4103

Project: North Point

RunNo.: LB132518

Analyte		Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: Corrosivity	ICV	pH	7.02	7	100	90-110	09/19/2024
Sample ID: Corrosivity	CCV1	pH	2.01	2.00	101	90-110	09/19/2024
Sample ID: Corrosivity	CCV2	pH	12.02	12.00	100	90-110	09/19/2024
Sample ID: Corrosivity	CCV3	pH	2.01	2.00	101	90-110	09/19/2024

Initial and Continuing Calibration Verification

Client: ENTACT

SDG No.: P4103

Project: North Point

RunNo.: LB132537

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV Hexavalent Chromium	mg/L	0.500	0.5	100	90-110	09/20/2024
Sample ID: CCV1 Hexavalent Chromium	mg/L	0.501	0.5	100	90-110	09/20/2024
Sample ID: CCV2 Hexavalent Chromium	mg/L	0.499	0.5	100	90-110	09/20/2024
Sample ID: CCV3 Hexavalent Chromium	mg/L	0.503	0.5	101	90-110	09/20/2024

Initial and Continuing Calibration Verification

Client: ENTACT

SDG No.: P4103

Project: North Point

RunNo.: LB132565

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1 ASTM Ammonia	mg/L	1	1	100	90-110	09/24/2024
Sample ID: CCV1 ASTM Ammonia	mg/L	0.99	1	99	90-110	09/24/2024
Sample ID: CCV2 ASTM Ammonia	mg/L	1.1	1	110	90-110	09/24/2024
Sample ID: CCV3 ASTM Ammonia	mg/L	1	1	100	90-110	09/24/2024

Initial and Continuing Calibration Verification

Client: ENTACT

SDG No.: P4103

Project: North Point

RunNo.: LB132579

Analyte		Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID:	ICV						
ASTM COD		mg/L	49.231	50	98	95-105	07/09/2024
Sample ID:	CCV1						
ASTM COD		mg/L	48.249	50	96	95-105	09/24/2024
Sample ID:	CCV2						
ASTM COD		mg/L	49.231	50	98	95-105	09/24/2024
Sample ID:	CCV3						
ASTM COD		mg/L	49.231	50	98	95-105	09/24/2024

Initial and Continuing Calibration Verification

Client: ENTACT

SDG No.: P4103

Project: North Point

RunNo.: LB132584

Analyte		Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID:	ICV1						
Cyanide		mg/L	0.098	0.099	99	90-110	09/24/2024
Sample ID:	CCV1						
Cyanide		mg/L	0.25	0.25	100	90-110	09/24/2024
Sample ID:	CCV2						
Cyanide		mg/L	0.25	0.25	100	90-110	09/24/2024
Sample ID:	CCV3						
Cyanide		mg/L	0.26	0.25	104	90-110	09/24/2024