

Initial and Continuing Calibration Blank Summary

Client: ENTACT

SDG No.: P4103

Project: North Point

RunNo.: LB132516

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: ICB1 Reactive Cyanide	mg/L	< 0.0025	0.0025	U	0.00099	0.005	09/19/2024
Sample ID: CCB1 Reactive Cyanide	mg/L	< 0.0025	0.0025	U	0.00099	0.005	09/19/2024
Sample ID: CCB2 Reactive Cyanide	mg/L	< 0.0025	0.0025	U	0.00099	0.005	09/19/2024
Sample ID: CCB3 Reactive Cyanide	mg/L	< 0.0025	0.0025	U	0.00099	0.005	09/19/2024
Sample ID: CCB4 Reactive Cyanide	mg/L	< 0.0025	0.0025	U	0.00099	0.005	09/19/2024

Initial and Continuing Calibration Blank Summary

Client: ENTACT

SDG No.: P4103

Project: North Point

RunNo.: LB132537

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: ICB Hexavalent Chromium	mg/L	< 0.0050	0.0050	U	0.0027	0.01	09/20/2024
Sample ID: CCB1 Hexavalent Chromium	mg/L	< 0.0050	0.0050	U	0.0027	0.01	09/20/2024
Sample ID: CCB2 Hexavalent Chromium	mg/L	< 0.0050	0.0050	U	0.0027	0.01	09/20/2024
Sample ID: CCB3 Hexavalent Chromium	mg/L	< 0.0050	0.0050	U	0.0027	0.01	09/20/2024

Initial and Continuing Calibration Blank Summary

Client: ENTACT

SDG No.: P4103

Project: North Point

RunNo.: LB132565

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: ICB1 ASTM Ammonia	mg/L	< 0.0500	0.0500	U	0.045	0.1	09/24/2024
Sample ID: CCB1 ASTM Ammonia	mg/L	< 0.0500	0.0500	U	0.045	0.1	09/24/2024
Sample ID: CCB2 ASTM Ammonia	mg/L	< 0.0500	0.0500	U	0.045	0.1	09/24/2024
Sample ID: CCB3 ASTM Ammonia	mg/L	< 0.0500	0.0500	U	0.045	0.1	09/24/2024

Initial and Continuing Calibration Blank Summary

Client: ENTACT

SDG No.: P4103

Project: North Point

RunNo.: LB132579

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: ICB ASTM COD	mg/L	< 5.0000	5.0000	U	2.35	10	07/09/2024
Sample ID: CCB1 ASTM COD	mg/L	< 5.0000	5.0000	U	2.35	10	09/24/2024
Sample ID: CCB2 ASTM COD	mg/L	< 5.0000	5.0000	U	2.35	10	09/24/2024
Sample ID: CCB3 ASTM COD	mg/L	< 5.0000	5.0000	U	2.35	10	09/24/2024

Initial and Continuing Calibration Blank Summary

Client: ENTACT

SDG No.: P4103

Project: North Point

RunNo.: LB132584

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: ICB1 Cyanide	mg/L	< 0.0025	0.0025	U	0.00099	0.005	09/24/2024
Sample ID: CCB1 Cyanide	mg/L	< 0.0025	0.0025	U	0.00099	0.005	09/24/2024
Sample ID: CCB2 Cyanide	mg/L	< 0.0025	0.0025	U	0.00099	0.005	09/24/2024
Sample ID: CCB3 Cyanide	mg/L	< 0.0025	0.0025	U	0.00099	0.005	09/24/2024