

Preparation Blank Summary

Client: ENTACT

SDG No.: P4103

Project: North Point

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: LB132520BL TS	%	< 2.5000	2.5000	U	1	5	09/19/2024
Sample ID: LB132521BL TVS	%	< 5.0000	5.0000	U	1	10	09/19/2024
Sample ID: LB132556BL ASTM TS	mg/L	< 2.5000	2.5000	U	1	5	09/23/2024
Sample ID: LB132564BL ASTM Oil and Grease	mg/L	< 2.5000	2.5000	U	0.4	5.0	09/24/2024
Sample ID: LB132579BL ASTM COD	mg/L	< 5.0000	5.0000	U	2.35	10.0	09/24/2024
Sample ID: PB163518BL Reactive Cyanide	mg/Kg	< 0.0250	0.0250	U	0.0088	0.05	09/19/2024
Sample ID: PB163530BL Hexavalent Chromium	mg/Kg	< 0.2000	0.2000	U	0.079	0.4	09/20/2024
Sample ID: PB163550BL Reactive Sulfide	mg/Kg	< 5.0000	5.0000	U	0.186	10	09/23/2024
Sample ID: PB163619BL Cyanide	mg/Kg	< 0.1250	0.1250	U	0.044	0.25	09/24/2024
Sample ID: PB163633BL ASTM Ammonia	mg/L	< 0.0500	0.0500	U	0.045	0.1	09/24/2024