

Instrument ID: CV1

Daily Analysis Runlog For Sequence/QC Batch ID # LB132531

Review By	Mohan	Review On	9/20/2024 8:18:44 PM
Supervise By	jaswal	Supervise On	9/20/2024 10:29:41 PM
STD. NAME	STD REF.#		
ICAL Standard	MP82394,MP82395,MP82396,MP82397,MP82398,MP82399		
ICV Standard	MP82400		
CCV Standard	MP82402		
ICSA Standard			
CRI Standard	MP82404		
LCS Standard			
Chk Standard	MP82407,MP82408		

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	S0	S0	CAL1	09/20/24 14:12		Mohan	OK
2	S0.2	S0.2	CAL2	09/20/24 14:14		Mohan	OK
3	S2.5	S2.5	CAL3	09/20/24 14:17		Mohan	OK
4	S5	S5	CAL4	09/20/24 14:19		Mohan	OK
5	S7.5	S7.5	CAL5	09/20/24 14:21		Mohan	OK
6	S10	S10	CAL6	09/20/24 14:24		Mohan	OK
7	ICV37	ICV37	ICV	09/20/24 14:28		Mohan	OK
8	ICB37	ICB37	ICB	09/20/24 14:30		Mohan	OK
9	CCV01	CCV01	CCV	09/20/24 14:32		Mohan	OK
10	CCB01	CCB01	CCB	09/20/24 14:34		Mohan	OK
11	CRA	CRA	CRDL	09/20/24 14:40		Mohan	OK
12	HighStd	HighStd	HIGH STD	09/20/24 14:42		Mohan	OK
13	ChkStd	ChkStd	SAM	09/20/24 14:44		Mohan	OK
14	PB163476BL	PB163476BL	MB	09/20/24 14:47		Mohan	OK
15	PB163476BS	PB163476BS	LCS	09/20/24 14:49		Mohan	OK
16	P4014-01	OR-03-09162024	SAM	09/20/24 14:52		Mohan	OK
17	P4023-01	TP-9	SAM	09/20/24 14:54		Mohan	OK
18	P4023-03	CONCRETE-TP-9	SAM	09/20/24 14:56		Mohan	OK

Instrument ID: CV1

Daily Analysis Runlog For Sequence/QC Batch ID # LB132531

Review By	Mohan	Review On	9/20/2024 8:18:44 PM
Supervise By	jaswal	Supervise On	9/20/2024 10:29:41 PM
STD. NAME	STD REF.#		
ICAL Standard	MP82394,MP82395,MP82396,MP82397,MP82398,MP82399		
ICV Standard	MP82400		
CCV Standard	MP82402		
ICSA Standard			
CRI Standard	MP82404		
LCS Standard			
Chk Standard	MP82407,MP82408		

19	P4027-01	DN-B-49	SAM	09/20/24 14:58		Mohan	OK
20	P4085-01	COMP-1	SAM	09/20/24 15:01		Mohan	OK
21	CCV02	CCV02	CCV	09/20/24 15:03		Mohan	OK
22	CCB02	CCB02	CCB	09/20/24 15:05		Mohan	OK
23	P4085-03	COMP-2	SAM	09/20/24 15:08		Mohan	OK
24	P4085-03DUP	COMP-2DUP	DUP	09/20/24 15:10		Mohan	OK
25	P4085-03MS	COMP-2MS	MS	09/20/24 15:15		Mohan	OK
26	P4085-03MSD	COMP-2MSD	MSD	09/20/24 15:17		Mohan	OK
27	P4086-01	SOIL	SAM	09/20/24 15:20		Mohan	OK
28	P4087-01	NB-614-COMP-01	SAM	09/20/24 15:22	High	Mohan	Dilution
29	P4087-07	NB-614-COMP-02	SAM	09/20/24 15:24		Mohan	OK
30	P4087-13	NB-614-COMP-03	SAM	09/20/24 15:27		Mohan	OK
31	P4087-19	NB-614-COMP-04	SAM	09/20/24 15:29		Mohan	OK
32	P4087-25	NB-614-COMP-05	SAM	09/20/24 15:31		Mohan	OK
33	CCV03	CCV03	CCV	09/20/24 15:34		Mohan	OK
34	CCB03	CCB03	CCB	09/20/24 15:36		Mohan	OK
35	P4087-31	NB-614-COMP-06	SAM	09/20/24 15:38		Mohan	OK
36	P4088-01	NB-614-COMP-07	SAM	09/20/24 15:40		Mohan	OK
37	P4088-07	NB-614-COMP-08	SAM	09/20/24 15:43		Mohan	OK
38	P4088-13	NB-614-COMP-09	SAM	09/20/24 15:45		Mohan	OK

Instrument ID: CV1

Daily Analysis Runlog For Sequence/QC Batch ID # LB132531

Review By	Mohan	Review On	9/20/2024 8:18:44 PM
Supervise By	jaswal	Supervise On	9/20/2024 10:29:41 PM
STD. NAME	STD REF.#		
ICAL Standard	MP82394,MP82395,MP82396,MP82397,MP82398,MP82399		
ICV Standard	MP82400		
CCV Standard	MP82402		
ICSA Standard			
CRI Standard	MP82404		
LCS Standard			
Chk Standard	MP82407,MP82408		

39	P4089-01	365-351-294	SAM	09/20/24 15:47		Mohan	OK
40	P4090-01	TR-05-091724	SAM	09/20/24 15:50		Mohan	OK
41	P4097-01	RT2299	SAM	09/20/24 15:52		Mohan	OK
42	P4098-01	HR-01-091724	SAM	09/20/24 15:54		Mohan	OK
43	P4098-03	HR-04-091724	SAM	09/20/24 15:57		Mohan	OK
44	PB163547BL	PB163547BL	MB	09/20/24 15:59		Mohan	OK
45	CCV04	CCV04	CCV	09/20/24 16:01		Mohan	OK
46	CCB04	CCB04	CCB	09/20/24 16:03		Mohan	OK
47	PB163547BS	PB163547BS	LCS	09/20/24 16:06		Mohan	OK
48	P4103-01	WC-22A	SAM	09/20/24 16:08		Mohan	OK
49	P4103-04	WC-14-5-7.5A	SAM	09/20/24 16:10		Mohan	OK
50	P4103-07	WC-14-5-7.5B	SAM	09/20/24 16:12		Mohan	OK
51	P4103-10	WC-14-7.5-10A	SAM	09/20/24 16:15		Mohan	OK
52	P4103-13	WC-14-7.5-10B	SAM	09/20/24 16:17		Mohan	OK
53	P4103-16	WC-23	SAM	09/20/24 16:19		Mohan	OK
54	P4120-13	RB24085	SAM	09/20/24 16:22		Mohan	OK
55	P4123-02	613-D	SAM	09/20/24 16:24		Mohan	OK
56	P4125-01	PL-01-09192024	SAM	09/20/24 16:26		Mohan	OK
57	CCV05	CCV05	CCV	09/20/24 16:28		Mohan	OK
58	CCB05	CCB05	CCB	09/20/24 16:31		Mohan	OK

Instrument ID: CV1

Daily Analysis Runlog For Sequence/QC Batch ID # LB132531

Review By	Mohan	Review On	9/20/2024 8:18:44 PM
Supervise By	jaswal	Supervise On	9/20/2024 10:29:41 PM

STD. NAME	STD REF.#
ICAL Standard	MP82394,MP82395,MP82396,MP82397,MP82398,MP82399
ICV Standard	MP82400
CCV Standard	MP82402
ICSA Standard	
CRI Standard	MP82404
LCS Standard	
Chk Standard	MP82407,MP82408

59	P4126-01	NB-08-09192024	SAM	09/20/24 16:33		Mohan	OK
60	P4126-01DUP	NB-08-09192024DUP	DUP	09/20/24 16:35		Mohan	OK
61	P4126-01MS	NB-08-09192024MS	MS	09/20/24 16:41		Mohan	OK
62	P4126-01MSD	NB-08-09192024MSD	MSD	09/20/24 16:44		Mohan	OK
63	CCV06	CCV06	CCV	09/20/24 16:46		Mohan	OK
64	CCB06	CCB06	CCB	09/20/24 16:48		Mohan	OK
65	P4085-03L	COMP-2L	SD	09/20/24 16:51		Mohan	OK
66	P4085-03A	COMP-2A	PS	09/20/24 16:53		Mohan	OK
67	P4126-01L	NB-08-09192024L	SD	09/20/24 16:55		Mohan	OK
68	P4126-01A	NB-08-09192024A	PS	09/20/24 16:57		Mohan	OK
69	P4087-01DL	NB-614-COMP-01DL	SAM	09/20/24 17:00	Report 2X	Mohan	Confirms
70	CCV07	CCV07	CCV	09/20/24 17:02		Mohan	OK
71	CCB07	CCB07	CCB	09/20/24 17:04		Mohan	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB132557

Review By	kareem	Review On	9/26/2024 11:58:17 AM
Supervise By		Supervise On	
STD. NAME	STD REF.#		
ICAL Standard	MP82441 MP82476 MP82477 MP82478 MP82479 MP82484		
ICV Standard	MP82485		
CCV Standard	MP82488		
ICSA Standard	MP82486 MP82487		
CRI Standard	MP82487		
LCS Standard			
Chk Standard	MP82491 MP82492		

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	S0	S0	CAL1	09/23/24 12:08		Kareem	OK
2	S1	S1	CAL2	09/23/24 12:13		Kareem	OK
3	S2	S2	CAL3	09/23/24 12:17		Kareem	OK
4	S3	S3	CAL4	09/23/24 12:21		Kareem	OK
5	S4	S4	CAL5	09/23/24 12:25		Kareem	OK
6	S5	S5	CAL6	09/23/24 12:30		Kareem	OK
7	ICV01	ICV01	ICV	09/23/24 12:49	ICV fail for Ag (200.7) (95-105)	Kareem	OK
8	LLICV01	LLICV01	LLICV	09/23/24 12:58		Kareem	OK
9	ICB01	ICB01	ICB	09/23/24 13:03		Kareem	OK
10	CRI01	CRI01	CRDL	09/23/24 13:07		Kareem	OK
11	ICSA01	ICSA01	ICSA	09/23/24 13:17		Kareem	OK
12	ICSAB01	ICSAB01	ICSAB	09/23/24 13:22		Kareem	OK
13	ICSADL	ICSADL	ICSA	09/23/24 13:29		Kareem	OK
14	ICSABDL	ICSABDL	ICSAB	09/23/24 13:33		Kareem	OK
15	CCV01	CCV01	CCV	09/23/24 13:38		Kareem	OK
16	CCB01	CCB01	CCB	09/23/24 13:43		Kareem	OK
17	P4126-01	NB-08-09192024	SAM	09/23/24 13:47		Kareem	OK
18	CCV02	CCV02	CCV	09/23/24 13:51		Kareem	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB132557

Review By	kareem	Review On	9/26/2024 11:58:17 AM
Supervise By		Supervise On	
STD. NAME	STD REF.#		
ICAL Standard	MP82441 MP82476 MP82477 MP82478 MP82479 MP82484		
ICV Standard	MP82485		
CCV Standard	MP82488		
ICSA Standard	MP82486 MP82487		
CRI Standard	MP82487		
LCS Standard			
Chk Standard	MP82491 MP82492		

19	CCB02	CCB02	CCB	09/23/24 13:55		Kareem	OK
20	LR1	LR1	HIGH STD	09/23/24 14:04		Kareem	OK
21	LR2	LR2	HIGH STD	09/23/24 14:09		Kareem	OK
22	P4103-01	WC-22A	SAM	09/23/24 14:14		Kareem	OK
23	P4103-04	WC-14-5-7.5A	SAM	09/23/24 14:19		Kareem	OK
24	P4103-07	WC-14-5-7.5B	SAM	09/23/24 14:23		Kareem	OK
25	P4103-10	WC-14-7.5-10A	SAM	09/23/24 14:27		Kareem	OK
26	P4103-13	WC-14-7.5-10B	SAM	09/23/24 14:31		Kareem	OK
27	P4103-16	WC-23	SAM	09/23/24 14:35		Kareem	OK
28	P4126-01DUP	NB-08-09192024DUP	DUP	09/23/24 14:39		Kareem	OK
29	P4126-01L	NB-08-09192024L	SD	09/23/24 14:43		Kareem	OK
30	CCV03	CCV03	CCV	09/23/24 14:47		Kareem	OK
31	CCB03	CCB03	CCB	09/23/24 14:52		Kareem	OK
32	P4126-01MS	NB-08-09192024MS	MS	09/23/24 14:56	0.1ml of m6001 and m6010 were added to 10ml of the sample	Kareem	OK
33	P4126-01MSD	NB-08-09192024MSD	MSD	09/23/24 15:00	0.1ml of m6001 and m6010 were added to 10ml of the sample	Kareem	OK
34	P4126-01A	NB-08-09192024A	PS	09/23/24 15:04	0.1ml of m6001 and m6010 were added to 10ml of the sample	Kareem	OK
35	P3956-02	EFF-WW	SAM	09/23/24 15:08		Kareem	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB132557

Review By	kareem	Review On	9/26/2024 11:58:17 AM
Supervise By		Supervise On	

STD. NAME	STD REF.#
ICAL Standard	MP82441 MP82476 MP82477 MP82478 MP82479 MP82484
ICV Standard	MP82485
CCV Standard	MP82488
ICSA Standard	MP82486 MP82487
CRI Standard	MP82487
LCS Standard	
Chk Standard	MP82491 MP82492

36	P3956-02DUP	EFF-WWDUP	DUP	09/23/24 15:13		Kareem	OK
37	P3956-02L	EFF-WWL	SD	09/23/24 15:17		Kareem	OK
38	P3956-02MS	EFF-WWMS	MS	09/23/24 15:21	0.1ml of m6001 and m6010 were added to 10ml of the sample	Kareem	OK
39	P3956-02MSD	EFF-WWMSD	MSD	09/23/24 15:25	0.1ml of m6001 and m6010 were added to 10ml of the sample	Kareem	OK
40	P3956-02A	EFF-WWA	PS	09/23/24 15:29	0.1ml of m6001 and m6010 were added to 10ml of the sample	Kareem	OK
41	PB163546BL	PB163546BL	MB	09/23/24 15:33		Kareem	OK
42	CCV04	CCV04	CCV	09/23/24 15:37		Kareem	OK
43	CCB04	CCB04	CCB	09/23/24 15:42		Kareem	OK
44	PB163546BS	PB163546BS	LCS	09/23/24 15:49	0.1ml of m6001 and m6010 were added to 10ml of the sample	Kareem	OK
45	P4103-02	WC-22A	SAM	09/23/24 15:53		Kareem	OK
46	P4103-05	WC-14-5-7.5A	SAM	09/23/24 15:57		Kareem	OK
47	P4103-08	WC-14-5-7.5B	SAM	09/23/24 16:01		Kareem	OK
48	P4103-11	WC-14-7.5-10A	SAM	09/23/24 16:06		Kareem	OK
49	P4103-04RE	WC-14-5-7.5ARE	SAM	09/23/24 16:10	NOT USE	Kareem	Not Ok
50	P4103-17	WC-23	SAM	09/23/24 16:15		Kareem	OK
51	P4120-02	T1	SAM	09/23/24 16:19		Kareem	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB132557

Review By	kareem	Review On	9/26/2024 11:58:17 AM
Supervise By		Supervise On	
STD. NAME	STD REF.#		
ICAL Standard	MP82441 MP82476 MP82477 MP82478 MP82479 MP82484		
ICV Standard	MP82485		
CCV Standard	MP82488		
ICSA Standard	MP82486 MP82487		
CRI Standard	MP82487		
LCS Standard			
Chk Standard	MP82491 MP82492		

52	P4120-04	T2	SAM	09/23/24 16:24		Kareem	OK
53	P4120-06	SF-6-A	SAM	09/23/24 16:28		Kareem	OK
54	CCV05	CCV05	CCV	09/23/24 16:33		Kareem	OK
55	CCB05	CCB05	CCB	09/23/24 16:37		Kareem	OK
56	P4120-08	SF-6-B	SAM	09/23/24 16:41		Kareem	OK
57	P4120-10	COMP-1	SAM	09/23/24 16:46		Kareem	OK
58	P4120-12	COMP-2	SAM	09/23/24 16:51		Kareem	OK
59	P4120-14	RB24085	SAM	09/23/24 16:55		Kareem	OK
60	P4121-01	COMP-1	SAM	09/23/24 16:59		Kareem	OK
61	P4121-02	COMP-2	SAM	09/23/24 17:04		Kareem	OK
62	P4123-01	910-B	SAM	09/23/24 17:08		Kareem	OK
63	P4123-01DUP	910-BDUP	DUP	09/23/24 17:13		Kareem	OK
64	P4123-01L	910-BL	SD	09/23/24 17:17		Kareem	OK
65	P4123-01MS	910-BMS	MS	09/23/24 17:21	0.1ml of m6001 and m6010 were added to 10ml of the sample	Kareem	OK
66	CCV06	CCV06	CCV	09/23/24 17:26		Kareem	OK
67	CCB06	CCB06	CCB	09/23/24 17:30		Kareem	OK
68	P4123-01MSD	910-BMSD	MSD	09/23/24 17:34	0.1ml of m6001 and m6010 were added to 10ml of the sample	Kareem	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB132557

Review By	kareem	Review On	9/26/2024 11:58:17 AM
Supervise By		Supervise On	

STD. NAME	STD REF.#
ICAL Standard	MP82441 MP82476 MP82477 MP82478 MP82479 MP82484
ICV Standard	MP82485
CCV Standard	MP82488
ICSA Standard	MP82486 MP82487
CRI Standard	MP82487
LCS Standard	
Chk Standard	MP82491 MP82492

69	P4123-01A	910-BA	PS	09/23/24 17:39	0.1ml of m6001 and m6010 were added to 10ml of the sample	Kareem	OK
70	PB163511TB	PB163511TB	MB	09/23/24 17:43		Kareem	OK
71	PB163580BL	PB163580BL	MB	09/23/24 17:56		Kareem	OK
72	PB163580BS	PB163580BS	LCS	09/23/24 18:00	0.1ml of m6001 and m6010 were added to 10ml of the sample	Kareem	OK
73	CCV07	CCV07	CCV	09/23/24 18:04		Kareem	OK
74	CCB07	CCB07	CCB	09/23/24 18:08		Kareem	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB132578

Review By	mohan	Review On	9/25/2024 4:26:44 AM
Supervise By	kareem	Supervise On	9/26/2024 12:16:35 PM
STD. NAME	STD REF.#		
ICAL Standard	MP82441 MP82476 MP82477 MP82478 MP82479 MP82484		
ICV Standard	MP82485		
CCV Standard	MP82488		
ICSA Standard	MP82486 MP82487		
CRI Standard	mp82484		
LCS Standard			
Chk Standard	MP82491 MP82492		

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	S0	S0	CAL1	09/24/24 12:51		Kareem	OK
2	S1	S1	CAL2	09/24/24 12:55		Kareem	OK
3	S2	S2	CAL3	09/24/24 13:00		Kareem	OK
4	S3	S3	CAL4	09/24/24 13:04		Kareem	OK
5	S4	S4	CAL5	09/24/24 13:08		Kareem	OK
6	S5	S5	CAL6	09/24/24 13:12		Kareem	OK
7	ICV01	ICV01	ICV	09/24/24 13:34		Kareem	OK
8	LLICV01	LLICV01	LLICV	09/24/24 13:38		Kareem	OK
9	ICB01	ICB01	ICB	09/24/24 13:56		Kareem	OK
10	CRI01	CRI01	CRDL	09/24/24 14:00		Kareem	OK
11	ICSA01	ICSA01	ICSA	09/24/24 14:04		Kareem	OK
12	ICSAB01	ICSAB01	ICSAB	09/24/24 14:14		Kareem	OK
13	CCV01	CCV01	CCV	09/24/24 14:25		Kareem	OK
14	CCB01	CCB01	CCB	09/24/24 14:29		Kareem	OK
15	P4138-02	LIBERTY-AVE-TP-CO	SAM	09/24/24 14:33		Kareem	OK
16	P4139-05	350	SAM	09/24/24 14:38		Kareem	OK
17	P4139-06	371	SAM	09/24/24 14:42		Kareem	OK
18	P4140-01	MOO-23-00305	SAM	09/24/24 14:47		Kareem	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB132578

Review By	mohan	Review On	9/25/2024 4:26:44 AM
Supervise By	kareem	Supervise On	9/26/2024 12:16:35 PM
STD. NAME	STD REF.#		
ICAL Standard	MP82441 MP82476 MP82477 MP82478 MP82479 MP82484		
ICV Standard	MP82485		
CCV Standard	MP82488		
ICSA Standard	MP82486 MP82487		
CRI Standard	mp82484		
LCS Standard			
Chk Standard	MP82491 MP82492		

19	P4144-02	1-3-SBG	SAM	09/24/24 14:51		Kareem	OK
20	P4144-04	4-6-SBG	SAM	09/24/24 14:55		Kareem	OK
21	P4144-04DUP	4-6-SBGDUP	DUP	09/24/24 15:00		Kareem	OK
22	P4144-04L	4-6-SBGL	SD	09/24/24 15:04		Kareem	OK
23	P4144-04MS	4-6-SBGMS	MS	09/24/24 15:09	0.1ml of m6010 and m6001 were added to 10ml of sample	Kareem	OK
24	P4144-04MSD	4-6-SBGMSD	MSD	09/24/24 15:13	0.1ml of m6010 and m6001 were added to 10ml of sample	Kareem	OK
25	CCV02	CCV02	CCV	09/24/24 15:17		Kareem	OK
26	CCB02	CCB02	CCB	09/24/24 15:21		Kareem	OK
27	P4144-04A	4-6-SBGA	PS	09/24/24 15:25	0.1ml of m6010 and m6001 were added to 10ml of sample	Kareem	OK
28	PB163591TB	PB163591TB	MB	09/24/24 15:30		Kareem	OK
29	P4123-04	910-A	SAM	09/24/24 15:34		Kareem	OK
30	P4140-02	306-307	SAM	09/24/24 15:38		Kareem	OK
31	PB163590TB	PB163590TB	MB	09/24/24 15:43		Kareem	OK
32	PB163623BL	PB163623BL	MB	09/24/24 15:47		Kareem	OK
33	PB163623BS	PB163623BS	LCS	09/24/24 15:52	0.1ml of m6010 and m6001 were added to 10ml of sample	Kareem	OK
34	P4138-01	LIBERTY-AVE-TP-CO	SAM	09/24/24 15:56		Kareem	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB132578

Review By	mohan	Review On	9/25/2024 4:26:44 AM
Supervise By	kareem	Supervise On	9/26/2024 12:16:35 PM

STD. NAME	STD REF.#
ICAL Standard	MP82441 MP82476 MP82477 MP82478 MP82479 MP82484
ICV Standard	MP82485
CCV Standard	MP82488
ICSA Standard	MP82486 MP82487
CRI Standard	mp82484
LCS Standard	
Chk Standard	MP82491 MP82492

35	P4138-01DUP	LIBERTY-AVE-TP-CO	DUP	09/24/24 16:00		Kareem	OK
36	P4138-01L	LIBERTY-AVE-TP-CO	SD	09/24/24 16:04		Kareem	OK
37	CCV03	CCV03	CCV	09/24/24 16:08		Kareem	OK
38	CCB03	CCB03	CCB	09/24/24 16:12		Kareem	OK
39	P4138-01MS	LIBERTY-AVE-TP-CO	MS	09/24/24 16:17	0.1ml of m6010 and m6001 were added to 10ml of sample	Kareem	OK
40	P4138-01MSD	LIBERTY-AVE-TP-CO	MSD	09/24/24 16:20	0.1ml of m6010 and m6001 were added to 10ml of sample	Kareem	OK
41	P4138-01A	LIBERTY-AVE-TP-CO	PS	09/24/24 16:24	0.1ml of m6010 and m6001 were added to 10ml of sample	Kareem	OK
42	P4139-01	350	SAM	09/24/24 16:28		Kareem	OK
43	P4139-02	371	SAM	09/24/24 16:33		Kareem	OK
44	P4142-01	OR-02-09202024	SAM	09/24/24 16:37		Kareem	OK
45	P4146-01	CL-02-092024	SAM	09/24/24 16:41		Kareem	OK
46	P4148-01	STOCK-PILE	SAM	09/24/24 16:45		Kareem	OK
47	P4149-13	COMP-6	SAM	09/24/24 16:49		Kareem	OK
48	P4149-14	COMP-7	SAM	09/24/24 16:54		Kareem	OK
49	CCV04	CCV04	CCV	09/24/24 16:58		Kareem	OK
50	CCB04	CCB04	CCB	09/24/24 17:02		Kareem	OK
51	P4103-14	WC-14-7.5-10B	SAM	09/24/24 17:07		Kareem	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB132578

Review By	mohan	Review On	9/25/2024 4:26:44 AM
Supervise By	kareem	Supervise On	9/26/2024 12:16:35 PM
STD. NAME	STD REF.#		
ICAL Standard	MP82441 MP82476 MP82477 MP82478 MP82479 MP82484		
ICV Standard	MP82485		
CCV Standard	MP82488		
ICSA Standard	MP82486 MP82487		
CRI Standard	mp82484		
LCS Standard			
Chk Standard	MP82491 MP82492		

52	PB163545BL	PB163545BL	MB	09/24/24 17:11		Kareem	OK
53	PB163545BS	PB163545BS	LCS	09/24/24 17:15	0.1ml of m6010 and m6001 were added to 10ml of sample	Kareem	OK
54	P3989-08	TP-3	SAM	09/24/24 17:19		Kareem	OK
55	P3989-08DUP	TP-3DUP	DUP	09/24/24 17:24		Kareem	OK
56	P3989-08L	TP-3L	SD	09/24/24 17:28		Kareem	OK
57	P3989-08MS	TP-3MS	MS	09/24/24 17:33	0.1ml of m6010 and m6001 were added to 10ml of sample	Kareem	OK
58	P3989-08MSD	TP-3MSD	MSD	09/24/24 17:37	0.1ml of m6010 and m6001 were added to 10ml of sample	Kareem	OK
59	P3989-08A	TP-3A	PS	09/24/24 17:41	0.1ml of m6010 and m6001 were added to 10ml of sample	Kareem	OK
60	P4149-15	COMP-8	SAM	09/24/24 17:45		Kareem	OK
61	CCV05	CCV05	CCV	09/24/24 17:49		Kareem	OK
62	CCB05	CCB05	CCB	09/24/24 17:54		Kareem	OK
63	LR1	LR1	HIGH STD	09/24/24 17:59		Kareem	OK
64	LR2	LR2	HIGH STD	09/24/24 18:04		Kareem	OK
65	PB163601BL	PB163601BL	MB	09/24/24 18:09		Kareem	OK
66	PB163601BS	PB163601BS	LCS	09/24/24 18:14	0.1ml of m6010 and m6001 were added to 10ml of sample	Kareem	OK
67	P4144-01DL	1-3-SBGDL	SAM	09/24/24 18:18	10X for Cr,Fe,Zn, Ag	Kareem	Confirms

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB132578

Review By	mohan	Review On	9/25/2024 4:26:44 AM
Supervise By	kareem	Supervise On	9/26/2024 12:16:35 PM
STD. NAME	STD REF.#		
ICAL Standard	MP82441 MP82476 MP82477 MP82478 MP82479 MP82484		
ICV Standard	MP82485		
CCV Standard	MP82488		
ICSA Standard	MP82486 MP82487		
CRI Standard	mp82484		
LCS Standard			
Chk Standard	MP82491 MP82492		

68	P4144-03DL	4-6-SBGDL	SAM	09/24/24 18:22	10X for Cr,Fe,Zn, Ag	Kareem	Confirms
69	CCV06	CCV06	CCV	09/24/24 18:26		Kareem	OK
70	CCB06	CCB06	CCB	09/24/24 18:38		Kareem	OK
71	P4144-01	1-3-SBG	SAM	09/24/24 18:42	Cr,Fe,Zn High , Ag Oversaturated	Kareem	Dilution
72	P4144-03	4-6-SBG	SAM	09/24/24 18:46	Cr,Fe,Zn High , Ag Oversaturated	Kareem	Dilution
73	CCV07	CCV07	CCV	09/24/24 18:50		Kareem	OK
74	CCB07	CCB07	CCB	09/24/24 18:55		Kareem	OK