


LB132543, ANALYSIS DATE:09-21-24, INSTRUMENT ID:CV1

7470A



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	-0.002	-0.002	38	0.000	38	0				2
0.2	0.200	0.204	0.004	741	0.0 %	741					3
2.5	2.500	2.586	0.086	8877	0.0 %	8877					-4
5.0	5.000	4.821	-0.179	16509	0.0 %	16509					2
7.5	7.500	7.614	0.114	26049	0.0 %	26049					0
10.0	10.000	9.977	-0.023	34118	0.0 %	34118					

LB132543, INSTRUMENT ID:CV1

Method: 7470A Operator: Admin Date of Analysis: 21 Sep 2024 20:55:16

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1		38	-	0.0000	7470A	PPB	21 Sep 2024 20:58:13	S	Std
0.2 - 1		741	-	0.2000	7470A	PPB	21 Sep 2024 21:00:30	S	Std
2.5 - 1		8877	-	2.5000	7470A	PPB	21 Sep 2024 21:02:47	S	Std
5.0 - 1		16509	-	5.0000	7470A	PPB	21 Sep 2024 21:05:04	S	Std
7.5 - 1		26049	-	7.5000	7470A	PPB	21 Sep 2024 21:07:22	S	Std
10.0 - 1		34118	-	10.0000	7470A	PPB	21 Sep 2024 21:12:51	S	Std
ICV01 - 1	ICV01	13908	4.0593	-	7470A	PPB	21 Sep 2024 21:15:40	U	SMPL
ICB01 - 1	ICB01	-151	-0.0574	-	7470A	PPB	21 Sep 2024 21:17:55	U	SMPL
CCV01 - 1	CCV01	18059	5.2747	-	7470A	PPB	21 Sep 2024 21:20:11	U	SMPL
CCB01 - 1	CCB01	-200	-0.0717	-	7470A	PPB	21 Sep 2024 21:22:26	U	SMPL
CRA - 1	CRA	622	0.1690	-	7470A	PPB	21 Sep 2024 21:24:43	U	SMPL
HighStd - 1	HighStd	33900	9.9131	-	7470A	PPB	21 Sep 2024 21:26:57	U	SMPL
ChkStd - 1	ChkStd	24844	7.2614	-	7470A	PPB	21 Sep 2024 21:29:12	U	SMPL
PB163584BL - 1	PBW	174	0.0378	-	7470A	PPB	21 Sep 2024 21:31:31	U	SMPL
PB163584BS - 1	LCSW	14409	4.2060	-	7470A	PPB	21 Sep 2024 21:33:49	U	SMPL
P4099-01 - 1	GAS-TRANSMISSION-DEBRIS	2405	0.6911	-	7470A	PPB	21 Sep 2024 21:36:04	U	SMPL
P4103-02 - 1	WC-22A	-284	-0.0963	-	7470A	PPB	21 Sep 2024 21:38:22	U	SMPL
P4103-05 - 1	WC-14-5-7.5A	14	-0.0091	-	7470A	PPB	21 Sep 2024 21:40:41	U	SMPL
P4103-08 - 1	WC-14-5-7.5B	50	0.0015	-	7470A	PPB	21 Sep 2024 21:42:57	U	SMPL
P4103-11 - 1	WC-14-7.5-10A	48	0.0009	-	7470A	PPB	21 Sep 2024 21:45:12	U	SMPL
CCV02 - 1	CCV02	17244	5.0361	-	7470A	PPB	21 Sep 2024 21:47:28	U	SMPL
CCB02 - 1	CCB02	-106	-0.0442	-	7470A	PPB	21 Sep 2024 21:49:43	U	SMPL
P4103-14 - 1	WC-14-7.5-10B	38	-0.0020	-	7470A	PPB	21 Sep 2024 21:52:00	U	SMPL
P4103-17 - 1	WC-23	27	-0.0052	-	7470A	PPB	21 Sep 2024 21:54:15	U	SMPL
P4120-14 - 1	RB24085	108	0.0185	-	7470A	PPB	21 Sep 2024 21:56:30	U	SMPL
P4121-01 - 1	COMP-1	13	-0.0093	-	7470A	PPB	21 Sep 2024 21:58:45	U	SMPL
P4121-02 - 1	COMP-2	43	-0.0006	-	7470A	PPB	21 Sep 2024 22:01:01	U	SMPL
P4123-01 - 1	910-B	68	0.0068	-	7470A	PPB	21 Sep 2024 22:03:19	U	SMPL
P4123-01DUP - 1	910-BDUP	73	0.0082	-	7470A	PPB	21 Sep 2024 22:05:34	U	SMPL
P4123-01MS - 1	910-BMS	14088	4.1120	-	7470A	PPB	21 Sep 2024 22:07:51	U	SMPL
P4123-01MSD - 1	910-BMSD	13099	3.8224	-	7470A	PPB	21 Sep 2024 22:10:08	U	SMPL
P4123-01A - 1	910-BA	13758	4.0153	-	7470A	PPB	21 Sep 2024 22:12:25	U	SMPL
CCV03 - 1	CCV03	16631	4.8566	-	7470A	PPB	21 Sep 2024 22:14:42	U	SMPL
CCB03 - 1	CCB03	61	0.0047	-	7470A	PPB	21 Sep 2024 22:16:59	U	SMPL
P412-01LX5 - 1		154	0.0319	-	7470A	PPB	21 Sep 2024 22:19:16	U	SMPL
PB163511TB - 1		68	0.0068	-	7470A	PPB	21 Sep 2024 22:21:31	U	SMPL
CCV04 - 1	CCV04	18018	5.2627	-	7470A	PPB	21 Sep 2024 22:23:47	U	SMPL
CCB04 - 1	CCB04	-109	-0.0451	-	7470A	PPB	21 Sep 2024 22:26:02	U	SMPL