

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: ENTACT **Level:** LOW **SDG No.:** P4103
Contract: ENTA05 **Lab Code:** CHEM **Case No.:** P4103 **SAS No.:** P4103
Matrix: Water **Sample ID:** P4123-01 **Client ID:** 910-BDUP
Percent Solids for Sample: NA **Duplicate ID** P4123-01DUP **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Aluminum	ug/L	20	3810		3680		3		P
Arsenic	ug/L	20	100	U	100	U			P
Barium	ug/L	20	2510		2530		1		P
Beryllium	ug/L	20	1.85	J	2.18	J	16		P
Cadmium	ug/L	20	1.26	J	1.56	J	21		P
Chromium	ug/L	20	7.51	J	7.41	J	1		P
Copper	ug/L	20	106		104		2		P
Iron	ug/L	20	2250		2330		3		P
Lead	ug/L	20	60.0	U	60.0	U			P
Mercury	ug/L	20	2.00	U	2.00	U			CV
Nickel	ug/L	20	26.2	J	26.6	J	2		P
Selenium	ug/L	20	100	U	100	U			P
Silver	ug/L	20	50.0	U	50.0	U			P
Zinc	ug/L	20	1930		1940		1		P

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: ENTACT **Level:** LOW **SDG No.:** P4103
Contract: ENTA05 **Lab Code:** CHEM **Case No.:** P4103 **SAS No.:** P4103
Matrix: Water **Sample ID:** P4123-01MS **Client ID:** 910-BMSD
Percent Solids for Sample: NA **Duplicate ID** P4123-01MSD **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Aluminum	ug/L	20	13400		13300		1		P
Arsenic	ug/L	20	3620		3490		4		P
Barium	ug/L	20	3530		3510		1		P
Beryllium	ug/L	20	881		1050		18		P
Cadmium	ug/L	20	889		880		1		P
Chromium	ug/L	20	1930		1960		2		P
Copper	ug/L	20	1530		1510		1		P
Iron	ug/L	20	16700		16900		1		P
Lead	ug/L	20	4390		4350		1		P
Mercury	ug/L	20	41.1		38.2		7		CV
Nickel	ug/L	20	2430		2400		1		P
Selenium	ug/L	20	9010		8690		4		P
Silver	ug/L	20	380		384		1		P
Zinc	ug/L	20	2970		2950		1		P