

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: ENTACT **SDG No.:** P4103
Contract: ENTA05 **Lab Code:** CHEM **Case No.:** P4103 **SAS No.:** P4103

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB163580BS							
Aluminum	ug/L	10000	9820		98	80 - 120	P
Arsenic	ug/L	4000	3720		93	80 - 120	P
Barium	ug/L	1000	999		100	80 - 120	P
Beryllium	ug/L	1000	1160		116	80 - 120	P
Cadmium	ug/L	1000	930		93	80 - 120	P
Chromium	ug/L	2000	1930		96	80 - 120	P
Copper	ug/L	1500	1570		105	80 - 120	P
Iron	ug/L	15000	13700		91	80 - 120	P
Lead	ug/L	5000	4730		95	80 - 120	P
Nickel	ug/L	2500	2530		101	80 - 120	P
Selenium	ug/L	10000	9530		95	80 - 120	P
Silver	ug/L	380	385		101	80 - 120	P
Zinc	ug/L	1000	1010		101	80 - 120	P

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: ENTACT

SDG No.: P4103

Contract: ENTA05

Lab Code: CHEM

Case No.: P4103

SAS No.: P4103

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB163584BS Mercury	ug/L	4.0	4.21		105	80 - 120	CV