

DIESEL RANGE ORGANICS INITIAL CALIBRATION SUMMARY

Lab Name: Chemtech Contract: ENTA05
 ProjectID: North Point
 Lab Code: CHEM Case No.: P4103 SAS No.: P4103 SDG No.: P4103

Calibration Sequence : FF091124		Test : Diesel Range Organics	
Concentration (PPM)	Area Count	Reference Factor	File ID
500	63216053	126432	FF014481.D
200	24873045	124365	FF014482.D
100	12144149	121441	FF014483.D
50	5328640	106573	FF014484.D
1000	108851924	108852	FF014485.D
AVG RF : 117533		% RSD : 7.805	AVG RT : 14.8996