

SOIL DIESEL RANGE ORGANICS LABORATORY CONTROL SPIKE/LABORATORY CONTROL SPIKE DUPLICATE RI

Lab Name: Chemtech **Client:** ENTACT
Lab Code: CHEM **Cas No:** P4103 **SAS No :** P4103 **SDG No:** P4103
Matrix Spike - EPA Sample No : PB163522BS **Datafile:** FF014539.D

COMPOUND	SPIKE ADDED ug/kg	CONCENTRATION ug/kg	LCS/LCSD CONCENTRATION ug/kg	% REC	QC LIMITS
DRO	6664	0	6799	102	68-131