

Continuing Calibration Report for SequenceID : FC091924AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FC067178.D

Aliphatic C9-C12	6866172.000	60.000	3.177	6.459	114436.200	133603.315	14.346
Aliphatic C12-C16	4543224.000	40.000	6.460	9.852	113580.600	128900.867	11.885
Aliphatic C16-C21	6493224.000	60.000	9.853	13.212	108220.400	121616.431	11.015
Aliphatic C21-C28	8085597.000	80.000	13.213	16.869	101069.963	111967.399	9.733
Aliphatic C28-C40	9967991.000	120.000	16.870	21.722	83066.592	88261.110	5.885
Aliphatic EPH	35956208.000	360.000	3.177	21.722	99878.356	111160.957	10.150