

Continuing Calibration Report for SequenceID : FC091924AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FC067193.D

Aliphatic C9-C12	6890097.000	60.000	3.177	6.459	114834.950	133603.315	14.048
Aliphatic C12-C16	4547338.000	40.000	6.460	9.852	113683.450	128900.867	11.806
Aliphatic C16-C21	6552052.000	60.000	9.853	13.212	109200.867	121616.431	10.209
Aliphatic C21-C28	8141290.000	80.000	13.213	16.869	101766.125	111967.399	9.111
Aliphatic C28-C40	10297528.000	120.000	16.870	21.722	85812.733	88261.110	2.774
Aliphatic EPH	36428305.000	360.000	3.177	21.722	101189.736	111160.957	8.970