

Continuing Calibration Report for SequenceID : FC092024AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FC067196.D

Aliphatic C9-C12	6841146.000	60.000			114019.100	133603.315	14.658
Aliphatic C12-C16	4498486.000	40.000			112462.150	128900.867	12.753
Aliphatic C16-C21	6462124.000	60.000			107702.067	121616.431	11.441
Aliphatic C21-C28	8099175.000	80.000			101239.688	111967.399	9.581
Aliphatic C28-C40	10350708.000	120.000			86255.900	88261.110	2.272
Aliphatic EPH	36251639.000	360.000			100698.997	111160.957	9.412