

Continuing Calibration Report for SequenceID : FC092024AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FC067203.D

Aliphatic C9-C12	8387529.000	60.000			139792.150	133603.315	-4.632
Aliphatic C12-C16	5515370.000	40.000			137884.250	128900.867	-6.969
Aliphatic C16-C21	7864204.000	60.000			131070.067	121616.431	-7.773
Aliphatic C21-C28	9613500.000	80.000			120168.750	111967.399	-7.325
Aliphatic C28-C40	11471389.000	120.000			95594.908	88261.110	-8.309
Aliphatic EPH	42851992.000	360.000			119033.311	111160.957	-7.082