

Continuing Calibration Report for SequenceID : FD091924AR

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FD048333.D

Aromatic C10-C12	7252716.000	40.000	4.108	5.827	181317.900	178623.258	-1.509
Aromatic C12-C16	10792977.000	60.000	5.828	8.434	179882.950	177158.638	-1.538
Aromatic C16-C21	13096353.000	80.000	8.435	12.696	163704.413	162091.063	-0.995
Aromatic C21-C36	26839639.000	180.000	12.697	18.099	149109.106	147829.005	-0.866
Aromatic EPH	57981685.000	360.000	4.108	18.099	161060.236	159308.207	-1.100