

Continuing Calibration Report for SequenceID : FD091924AR

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FD048348.D

Aromatic C10-C12	7520088.000	40.000	4.108	5.827	188002.200	178623.258	-5.251
Aromatic C12-C16	11374106.000	60.000	5.828	8.434	189568.433	177158.638	-7.005
Aromatic C16-C21	13900346.000	80.000	8.435	12.696	173754.325	162091.063	-7.195
Aromatic C21-C36	30318274.000	180.000	12.697	18.099	168434.856	147829.005	-13.939
Aromatic EPH	63112814.000	360.000	4.108	18.099	175313.372	159308.207	-10.047