

Continuing Calibration Report for SequenceID : FE091924AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FE050216.D

Aliphatic C9-C12	7620239.000	60.000	3.159	6.778	127003.983	137280.139	7.486
Aliphatic C12-C16	5204646.000	40.000	6.779	10.211	130116.150	137537.275	5.396
Aliphatic C16-C21	7737573.000	60.000	10.212	13.571	128959.550	135214.517	4.626
Aliphatic C21-C28	10015787.000	80.000	13.572	17.227	125197.338	129389.049	3.240
Aliphatic C28-C40	14702751.000	120.000	17.228	22.082	122522.925	120097.177	-2.020
Aliphatic EPH	45280996.000	360.000	3.159	22.082	125780.544	129483.210	2.860