

Continuing Calibration Report for SequenceID : FE091924AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FE050231.D

Aliphatic C9-C12	7569177.000	60.000	3.159	6.778	126152.950	137280.139	8.105
Aliphatic C12-C16	5168333.000	40.000	6.779	10.211	129208.325	137537.275	6.056
Aliphatic C16-C21	7734358.000	60.000	10.212	13.571	128905.967	135214.517	4.666
Aliphatic C21-C28	10107099.000	80.000	13.572	17.227	126338.738	129389.049	2.357
Aliphatic C28-C40	14863027.000	120.000	17.228	22.082	123858.558	120097.177	-3.132
Aliphatic EPH	45441994.000	360.000	3.159	22.082	126227.761	129483.210	2.514