

Continuing Calibration Report for SequenceID : FE092024AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FE050243.D

Aliphatic C9-C12	7900139.000	60.000	3.158	6.776	131668.983	137280.139	4.087
Aliphatic C12-C16	5326045.000	40.000	6.777	10.209	133151.125	137537.275	3.189
Aliphatic C16-C21	7884539.000	60.000	10.210	13.569	131408.983	135214.517	2.814
Aliphatic C21-C28	10451271.000	80.000	13.570	17.223	130640.888	129389.049	-0.967
Aliphatic C28-C40	15221268.000	120.000	17.224	22.077	126843.900	120097.177	-5.618
Aliphatic EPH	46783262.000	360.000	3.158	22.077	129953.506	129483.210	-0.363