

Continuing Calibration Report for SequenceID : FG091924AR

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FG014280.D

Aromatic C10-C12	4811465.000	40.000	4.376	6.245	120286.625	127441.161	5.614
Aromatic C12-C16	7596272.000	60.000	6.246	8.944	126604.533	131942.474	4.046
Aromatic C16-C21	9769532.000	80.000	8.945	13.266	122119.150	129947.021	6.024
Aromatic C21-C36	21525714.000	180.000	13.267	18.723	119587.300	126096.133	5.162
Aromatic EPH	43702983.000	360.000	4.376	18.723	121397.175	128075.724	5.215