

Continuing Calibration Report for SequenceID : FG091924AR

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FG014295.D

Aromatic C10-C12	5050678.000	40.000	4.376	6.245	126266.950	127441.161	0.921
Aromatic C12-C16	7870333.000	60.000	6.246	8.944	131172.217	131942.474	0.584
Aromatic C16-C21	10295353.000	80.000	8.945	13.266	128691.913	129947.021	0.966
Aromatic C21-C36	22517646.000	180.000	13.267	18.723	125098.033	126096.133	0.792
Aromatic EPH	45734010.000	360.000	4.376	18.723	127038.917	128075.724	0.810