

Continuing Calibration Report for SequenceID : FG092024AR

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FG014298.D

Aromatic C10-C12	5160532.000	40.000	4.374	6.243	129013.300	127441.161	-1.234
Aromatic C12-C16	8046436.000	60.000	6.244	8.941	134107.267	131942.474	-1.641
Aromatic C16-C21	10354206.000	80.000	8.942	13.263	129427.575	129947.021	0.400
Aromatic C21-C36	22486000.000	180.000	13.264	18.721	124922.222	126096.133	0.931
Aromatic EPH	46047174.000	360.000	4.374	18.721	127908.817	128075.724	0.130