

Continuing Calibration Report for SequenceID : FG092024AR

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FG014306.D

Aromatic C10-C12	5373548.000	40.000	4.374	6.243	134338.700	127441.161	-5.412
Aromatic C12-C16	8402648.000	60.000	6.244	8.941	140044.133	131942.474	-6.140
Aromatic C16-C21	10728828.000	80.000	8.942	13.263	134110.350	129947.021	-3.204
Aromatic C21-C36	23444214.000	180.000	13.264	18.721	130245.633	126096.133	-3.291
Aromatic EPH	47949238.000	360.000	4.374	18.721	133192.328	128075.724	-3.995