

SOLID EPH LABORATORY CONTROL SPIKE/LABORATORY CONTROL SPIKE DUPLICATE RECOVERY

Lab Name: Chemtech **Client:** ENTACT
Lab Code: CHEM **Cas No:** P4103 **SAS No :** P4103 **SDG No:** P4103
Sample No : PB163524BSD **Datafile:** FC067181.D
Client ID : PB163524BSD

COMPOUND	SPIKE ADDED mg/kg	LCS/LCSD CONCENTRATION mg/kg	% REC	Qual	RPD QC LIMITS	QC Limit Of RPD
Aliphatic C9-C12	10	6.15	62		2.9 (40-140)	50
Aliphatic C12-C16	6.7	5.41	81		2.2 (40-140)	50
Aliphatic C16-C21	10	9.65	97		1 (40-140)	50
Aliphatic C21-C28	13.3	13.4	100		2.3 (40-140)	50
Aliphatic C28-C40	20.0	23.0	115		3.6 (40-140)	50