

SOLID EPH LABORATORY CONTROL SPIKE/LABORATORY CONTROL SPIKE DUPLICATE RECOVERY

Lab Name: Chemtech **Client:** ENTACT
Lab Code: CHEM **Cas No:** P4103 **SAS No :** P4103 **SDG No:** P4103
Sample No : PB163524BS **Datafile:** FD048335.D
Client ID : PB163524BS

COMPOUND	SPIKE ADDED mg/kg	LCS/LCSD CONCENTRATION mg/kg	% REC	Qual	QC LIMITS
Aromatic C10-C12	6.7	5.50	82		(40-140)
Aromatic C12-C16	10.0	9.16	92		(40-140)
Aromatic C16-C21	16.7	18.1	108		(40-140)
Aromatic C21-C36	26.7	27.8	104		(40-140)